

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062910.D
 Acq On : 13 Aug 2020 14:50
 Operator : JC/MD
 Sample : L3619-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 14 05:57:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	136754	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	236445	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	216271	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	80055	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	117070	51.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.54%	
35) Dibromofluoromethane	7.55	113	82702	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
50) Toluene-d8	10.06	98	293171	46.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.86%	
62) 4-Bromofluorobenzene	12.38	95	103649	43.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.88%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062910.D
Acq On : 13 Aug 2020 14:50
Operator : JC/MD
Sample : L3619-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Aug 14 05:57:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

