

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062914.D
 Acq On : 13 Aug 2020 16:28
 Operator : JC/MD
 Sample : L3667-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S39-0205

Quant Time: Aug 14 06:11:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	138438	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	238623	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	241221	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88921	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	115743	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
35) Dibromofluoromethane	7.55	113	83941	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
50) Toluene-d8	10.06	98	311158	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
62) 4-Bromofluorobenzene	12.38	95	112775	46.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.66%	
Target Compounds						
64) Tetrachloroethene	10.61	164	1534	0.986	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062914.D
Acq On : 13 Aug 2020 16:28
Operator : JC/MD
Sample : L3667-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S39-0205

Quant Time: Aug 14 06:11:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

