

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062915.D
 Acq On : 13 Aug 2020 16:52
 Operator : JC/MD
 Sample : L3667-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 S39-ER-2020-1

Manual Integrations
 APPROVED

MMDadoda
 8/14/2020 4:49:50 PM

Quant Time: Aug 14 08:51:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	146622	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	239439	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	240117	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	92704	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	114804	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
35) Dibromofluoromethane	7.55	113	85766	52.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.72%	
50) Toluene-d8	10.06	98	312127	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
62) 4-Bromofluorobenzene	12.38	95	114639	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
Target Compounds						
16) Acetone	3.78	43	6375m	7.276	ug/l	Qvalue
25) 2-Butanone	6.80	43	8356	6.734	ug/l	96
29) Tetrahydrofuran	7.17	42	133681	165.827	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062915.D
Acq On : 13 Aug 2020 16:52
Operator : JC/MD
Sample : L3667-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
S39-ER-2020-1

Manual Integrations
APPROVED

MMDadoda
8/14/2020 4:49:50 PM

Quant Time: Aug 14 08:51:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

