

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062916.D
 Acq On : 13 Aug 2020 17:41
 Operator : JC/MD
 Sample : PB130936TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130936TB

Quant Time: Aug 14 08:51:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	129245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	219478	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	232167	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	86493	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108362	50.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.40%	
35) Dibromofluoromethane	7.55	113	75419	50.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.46%	
50) Toluene-d8	10.06	98	298010	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.38	95	111487	50.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.66%	
Target Compounds						
16) Acetone	3.78	43	21601	27.970	ug/l	92
20) Methylene Chloride	4.50	84	13450	8.891	ug/l	93
43) Isopropyl Acetate	8.13	43	40943	10.996	ug/l	98
51) 4-Methyl-2-Pentanone	9.95	43	5593	2.322	ug/l	90
95) Naphthalene	15.11	128	8484	5.684	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062916.D
Acq On : 13 Aug 2020 17:41
Operator : JC/MD
Sample : PB130936TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130936TB

Quant Time: Aug 14 08:51:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

