

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062920.D
 Acq On : 13 Aug 2020 19:18
 Operator : JC/MD
 Sample : PB130936ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130936ZHE#03

Quant Time: Aug 14 06:39:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	135867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	244011	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	258679	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	90714	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	115291	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
35) Dibromofluoromethane	7.55	113	84656	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
50) Toluene-d8	10.06	98	325719	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	12.38	95	116233	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
Target Compounds						
16) Acetone	3.78	43	20648	25.433	ug/l	92
20) Methylene Chloride	4.50	84	13161	8.223	ug/l	93
25) 2-Butanone	6.80	43	4420	3.844	ug/l	93
43) Isopropyl Acetate	8.13	43	44622	10.779	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062920.D
Acq On : 13 Aug 2020 19:18
Operator : JC/MD
Sample : PB130936ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130936ZHE#03

Quant Time: Aug 14 06:39:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

