

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081320\
 Data File : VN062923.D
 Acq On : 13 Aug 2020 20:31
 Operator : JC/MD
 Sample : L3508-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-081220

Quant Time: Aug 14 06:45:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	117775	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	218782	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	212958	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	77368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105248	54.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.08%	
35) Dibromofluoromethane	7.55	113	71776	47.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.92%	
50) Toluene-d8	10.06	98	274174	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
62) 4-Bromofluorobenzene	12.38	95	98737	44.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.44%	
Target Compounds						
16) Acetone	3.78	43	6796	9.657	ug/l	89
20) Methylene Chloride	4.51	84	3212	1.760	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081320\
Data File : VN062923.D
Acq On : 13 Aug 2020 20:31
Operator : JC/MD
Sample : L3508-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-081220

Quant Time: Aug 14 06:45:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

