

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081320\
 Data File : VN062925.D
 Acq On : 13 Aug 2020 21:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 06:49:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	0.716	0.602	15.9	81	0.00
3 P	Chloromethane	0.676	0.527	22.0	82	0.00
4 C	Vinyl Chloride	0.634	0.540	14.8#	86	0.00
5 T	Bromomethane	0.419	0.494	-17.9	129	0.00
6 T	Chloroethane	0.417	0.559	-34.1#	138	0.00
7 T	Trichlorofluoromethane	1.130	1.097	2.9	102	0.00
8 T	Diethyl Ether	0.323	0.272	15.8	87	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.409	18.5	91	0.00
10 T	Methyl Iodide	0.548	0.519	5.3	92	0.00
11 T	Tert butyl alcohol	0.136	0.111	18.4	82	0.00
12 CM	1,1-Dichloroethene	0.476	0.375	21.2#	88	0.00
13 T	Acrolein	0.082	0.087	-6.1	112	0.00
14 T	Allyl chloride	0.875	0.654	25.3#	82	0.00
15 T	Acrylonitrile	0.314	0.268	14.6	84	0.00
16 T	Acetone	0.299	0.231	22.7	83	0.00
17 T	Carbon Disulfide	1.404	0.945	32.7#	80	0.00
18 T	Methyl Acetate	0.794	0.629	20.8	88	0.00
19 T	Methyl tert-butyl Ether	1.937	1.743	10.0	90	0.00
20 T	Methylene Chloride	0.609	0.494	18.9	92	0.00
21 T	trans-1,2-Dichloroethene	0.500	0.408	18.4	87	0.00
22 T	Diisopropyl ether	1.864	1.582	15.1	85	0.00
23 T	Vinyl Acetate	1.417	1.114	21.4	82	0.00
24 P	1,1-Dichloroethane	1.092	0.898	17.8	88	0.00
25 T	2-Butanone	0.423	0.350	17.3	82	0.00
26 T	2,2-Dichloropropane	1.007	0.688	31.7#	74	0.00
27 T	cis-1,2-Dichloroethene	0.600	0.511	14.8	91	0.00
28 T	Bromochloromethane	0.527	0.442	16.1	90	0.00
29 T	Tetrahydrofuran	0.275	0.230	16.4	81	0.00
30 C	Chloroform	1.161	1.001	13.8#	92	0.00
31 T	Cyclohexane	1.002	0.782	22.0	83	0.00
32 T	1,1,1-Trichloroethane	1.038	0.897	13.6	89	0.00
33 S	1,2-Dichloroethane-d4	0.827	0.731	11.6	94	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.342	0.306	10.5	96	0.00
36 T	1,1-Dichloropropene	0.487	0.399	18.1	86	0.00
37 T	Ethyl Acetate	0.509	0.401	21.2	82	0.00
38 T	Carbon Tetrachloride	0.576	0.490	14.9	93	0.00
39 T	Methylcyclohexane	0.586	0.474	19.1	85	0.00
40 TM	Benzene	1.453	1.192	18.0	89	0.00
41 T	Methacrylonitrile	0.229	0.189	17.5	77	0.00
42 TM	1,2-Dichloroethane	0.618	0.525	15.0	90	0.00
43 T	Isopropyl Acetate	0.848	0.677	20.2	83	0.00
44 TM	Trichloroethene	0.354	0.295	16.7	91	0.00
45 C	1,2-Dichloropropane	0.393	0.327	16.8#	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081320\
 Data File : VN062925.D
 Acq On : 13 Aug 2020 21:20
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 06:49:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.258	0.222	14.0	90	0.00
47 T	Bromodichloromethane	0.579	0.520	10.2	92	0.00
48 T	Methyl methacrylate	0.417	0.337	19.2	83	0.00
49 T	1,4-Dioxane	0.008	0.007	12.5	86	0.00
50 S	Toluene-d8	1.335	1.295	3.0	103	0.00
51 T	4-Methyl-2-Pentanone	0.549	0.538	2.0	98	0.00
52 CM	Toluene	0.897	0.854	4.8#	99	0.00
53 T	t-1,3-Dichloropropene	0.636	0.588	7.5	94	0.00
54 T	cis-1,3-Dichloropropene	0.639	0.551	13.8	89	0.00
55 T	1,1,2-Trichloroethane	0.383	0.375	2.1	106	0.00
56 T	Ethyl methacrylate	0.577	0.584	-1.2	98	0.00
57 T	1,3-Dichloropropane	0.651	0.645	0.9	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.163	0.128	21.5	80	0.00
59 T	2-Hexanone	0.409	0.403	1.5	96	0.00
60 T	Dibromochloromethane	0.437	0.429	1.8	101	0.00
61 T	1,2-Dibromoethane	0.378	0.370	2.1	102	0.00
62 S	4-Bromofluorobenzene	0.505	0.490	3.0	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	116	0.00
64 T	Tetrachloroethene	0.323	0.270	16.4	100	0.00
65 PM	Chlorobenzene	1.045	0.894	14.4	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.420	0.370	11.9	102	0.00
67 C	Ethyl Benzene	1.870	1.631	12.8#	99	0.00
68 T	m/p-Xylenes	0.679	0.607	10.6	99	0.00
69 T	o-Xylene	0.636	0.592	6.9	101	0.00
70 T	Styrene	1.118	1.037	7.2	99	0.00
71 P	Bromoform	0.343	0.293	14.6	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
73 T	Isopropylbenzene	3.765	3.317	11.9	97	0.00
74 T	N-amyl acetate	1.526	1.379	9.6	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.325	1.123	15.2	100	0.00
76 T	1,2,3-Trichloropropane	1.348	1.024	24.0	88	0.00
77 T	Bromobenzene	0.935	0.800	14.4	100	0.00
78 T	n-propylbenzene	4.420	3.847	13.0	96	0.00
79 T	2-Chlorotoluene	2.771	2.352	15.1	96	0.00
80 T	1,3,5-Trimethylbenzene	3.266	2.980	8.8	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.442	0.360	18.6	87	0.00
82 T	4-Chlorotoluene	2.856	2.516	11.9	98	0.00
83 T	tert-Butylbenzene	2.764	2.498	9.6	99	0.00
84 T	1,2,4-Trimethylbenzene	3.125	2.945	5.8	99	0.00
85 T	sec-Butylbenzene	3.629	3.244	10.6	97	0.00
86 T	p-Isopropyltoluene	3.222	2.913	9.6	96	0.00
87 T	1,3-Dichlorobenzene	1.671	1.444	13.6	96	0.00
88 T	1,4-Dichlorobenzene	1.676	1.415	15.6	97	0.00
89 T	n-Butylbenzene	2.725	2.283	16.2	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081320\
Data File : VN062925.D
Acq On : 13 Aug 2020 21:20
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 06:49:20 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.694	0.557	19.7	94	0.00
91 T	1,2-Dichlorobenzene	1.567	1.382	11.8	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.285	0.236	17.2	90	0.00
93 T	1,2,4-Trichlorobenzene	0.734	0.591	19.5	86	0.00
94 T	Hexachlorobutadiene	0.545	0.416	23.7	90	0.00
95 T	Naphthalene	1.932	1.585	18.0	86	0.00
96 T	1,2,3-Trichlorobenzene	0.714	0.605	15.3	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6