

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
 Data File : VN073871.D
 Acq On : 13 Aug 2022 16:25
 Operator : JC\MD
 Sample : N4156-04DL 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FMC-0074DL

Quant Time: Aug 14 02:35:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	612468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	872482	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	807936	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	412144	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	306029	41.001	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	82.000%
35) Dibromofluoromethane	7.951	113	295174	53.877	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.760%
50) Toluene-d8	10.380	98	996506	46.222	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.440%
62) 4-Bromofluorobenzene	12.674	95	382277	50.482	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.960%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	24916	6.215	ug/l	100
25) 2-Butanone	7.269	43	12151	1.952	ug/l	99
52) Toluene	10.445	92	49902	3.222	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
Data File : VN073871.D
Acq On : 13 Aug 2022 16:25
Operator : JC\MD
Sample : N4156-04DL 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FMC-0074DL

Quant Time: Aug 14 02:35:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

