

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
 Data File : VN073877.D
 Acq On : 13 Aug 2022 18:53
 Operator : JC\MD
 Sample : PB146976ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146976ZHE#03

Quant Time: Aug 14 02:38:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.022	168	672097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	969583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	929143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	518476	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	306435	37.413	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	74.820%
35) Dibromofluoromethane	7.951	113	303095	49.783	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.560%
50) Toluene-d8	10.380	98	1020576	42.598	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	85.200%#
62) 4-Bromofluorobenzene	12.674	95	440374	52.230	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.460%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	68140	15.489	ug/l	96
18) Methyl Acetate	4.787	43	17136	1.566	ug/l	95
20) Methylene Chloride	5.016	84	32933	2.535	ug/l	97
37) Ethyl Acetate	7.339	43	84743	7.416	ug/l	97
43) Isopropyl Acetate	8.498	43	521970	33.816	ug/l #	82
52) Toluene	10.445	92	101337	5.888	ug/l	98
68) m/p-Xylenes	11.886	106	17073	1.344	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
Data File : VN073877.D
Acq On : 13 Aug 2022 18:53
Operator : JC\MD
Sample : PB146976ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146976ZHE#03

Quant Time: Aug 14 02:38:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

