

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
 Data File : VN073881.D
 Acq On : 13 Aug 2022 20:31
 Operator : JC\MD
 Sample : PB146976ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB146976ZHE#07

Quant Time: Aug 14 02:39:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 05:09:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	640612	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	919559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	876805	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.616	152	459415	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	288045	36.896	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	73.800%#
35) Dibromofluoromethane	7.951	113	288025	49.881	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.760%
50) Toluene-d8	10.380	98	964141	42.431	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.860%#
62) 4-Bromofluorobenzene	12.674	95	393346	49.350	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.700%
Target Compounds						
					Qvalue	
16) Acetone	4.222	43	94028	22.424	ug/l	93
18) Methyl Acetate	4.781	43	12247	1.174	ug/l	98
20) Methylene Chloride	5.022	84	24475	1.469	ug/l	89
37) Ethyl Acetate	7.340	43	76585	7.066	ug/l	98
43) Isopropyl Acetate	8.492	43	435383	29.741	ug/l #	85
52) Toluene	10.445	92	26542	1.626	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
Data File : VN073881.D
Acq On : 13 Aug 2022 20:31
Operator : JC\MD
Sample : PB146976ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB146976ZHE#07

Quant Time: Aug 14 02:39:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 05:09:32 2022
Response via : Initial Calibration

