



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081322\
Data File : VN073866.D
Acq On : 13 Aug 2022 11:05
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N072822W.M
Title : SW846 8260
Last Update : Wed Aug 10 05:09:32 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1824

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.4	10661	PASS
75	95	30	60	50.3	30741	PASS
95	95	100	100	100.0	61107	PASS
96	95	5	9	7.6	4648	PASS
173	174	0.00	2	1.1	678	PASS
174	95	50	100	97.1	59365	PASS
175	174	5	9	7.5	4464	PASS
176	174	95	101	95.7	56840	PASS
177	176	5	9	6.8	3880	PASS