

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
 Data File : VN087538.D
 Acq On : 13 Aug 2025 16:21
 Operator : JC\MD
 Sample : Q2732-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-A7-01-G

Quant Time: Aug 14 04:03:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	239708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	533270	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	494800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	227638	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	239523	58.889	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	117.780%	
35) Dibromofluoromethane	8.153	113	161141	43.806	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	87.620%	
50) Toluene-d8	10.547	98	624414	47.587	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	95.180%	
62) 4-Bromofluorobenzene	12.829	95	228200	47.073	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	292723	153.495	ug/l	97
20) Methylene Chloride	5.271	84	14756	3.988	ug/l #	83
25) 2-Butanone	7.477	43	36044	12.232	ug/l	97
52) Toluene	10.618	92	12182	1.276	ug/l	86
95) Naphthalene	15.617	128	42077	2.927	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
Data File : VN087538.D
Acq On : 13 Aug 2025 16:21
Operator : JC\MD
Sample : Q2732-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-A7-01-G

Quant Time: Aug 14 04:03:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

