

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
 Data File : VN087545.D
 Acq On : 13 Aug 2025 18:56
 Operator : JC\MD
 Sample : Q2836-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-A2-16-G

Quant Time: Aug 14 04:06:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	266901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	539961	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	488726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	223776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	268865	59.369	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.740%
35) Dibromofluoromethane	8.147	113	364	0.098	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.200%#
50) Toluene-d8	10.547	98	645476	48.582	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.829	95	240812	49.059	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.120%
Target Compounds						
						Qvalue
16) Acetone	4.424	43	282383	132.986	ug/l	95
20) Methylene Chloride	5.277	84	13733	3.227	ug/l #	88
25) 2-Butanone	7.477	43	29686	9.048	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
Data File : VN087545.D
Acq On : 13 Aug 2025 18:56
Operator : JC\MD
Sample : Q2836-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-A2-16-G

Quant Time: Aug 14 04:06:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

