

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
 Data File : VN087548.D
 Acq On : 13 Aug 2025 20:02
 Operator : JC\MD
 Sample : Q2838-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-11

Quant Time: Aug 14 04:08:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	210108	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.082	114	434746	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	407444	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	183460	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	236754	66.409	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 132.820%#		
35) Dibromofluoromethane	8.153	113	156353	52.137	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.280%		
50) Toluene-d8	10.547	98	549864	51.402	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.800%		
62) 4-Bromofluorobenzene	12.829	95	205059	51.885	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 103.780%		
Target Compounds						
						Qvalue
16) Acetone	4.424	43	74565	44.608	ug/l	95
18) Methyl Acetate	5.030	43	48119	11.458	ug/l	100
20) Methylene Chloride	5.265	84	23442	7.755	ug/l	88
25) 2-Butanone	7.483	43	13522	5.236	ug/l	# 74
37) Ethyl Acetate	7.553	43	33260	5.813	ug/l	99
43) Isopropyl Acetate	8.677	43	21462	2.416	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
Data File : VN087548.D
Acq On : 13 Aug 2025 20:02
Operator : JC\MD
Sample : Q2838-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-11

Quant Time: Aug 14 04:08:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

