

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
 Data File : VN087549.D
 Acq On : 13 Aug 2025 20:24
 Operator : JC\MD
 Sample : Q2838-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10

Quant Time: Aug 14 04:08:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 17 02:56:13 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	192055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.088	114	400540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.847	117	367362	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	165090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.565	65	214880	65.939	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 131.880%#			
35) Dibromofluoromethane	8.147	113	140276	50.771	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.540%			
50) Toluene-d8	10.547	98	493980	50.121	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.240%			
62) 4-Bromofluorobenzene	12.829	95	181798	49.928	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.860%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	72761	47.620	ug/l	91
18) Methyl Acetate	5.018	43	48819	12.717	ug/l	99
20) Methylene Chloride	5.265	84	20531	7.403	ug/l #	92
25) 2-Butanone	7.483	43	14369	6.086	ug/l	93
37) Ethyl Acetate	7.553	43	30518	5.789	ug/l	99
43) Isopropyl Acetate	8.677	43	19152	2.340	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081325\
Data File : VN087549.D
Acq On : 13 Aug 2025 20:24
Operator : JC\MD
Sample : Q2838-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10

Quant Time: Aug 14 04:08:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071625W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 17 02:56:13 2025
Response via : Initial Calibration

