

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081418\
 Data File : VN050613.D
 Acq On : 14 Aug 2018 19:43
 Operator : MD\SY
 Sample : J4469-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 960-IW-21(17)

Quant Time: Aug 15 07:11:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	670176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1018571	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	878169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	298973	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	425202	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
35) Dibromofluoromethane	7.59	113	407961	50.17	ug/l	0.00
Spiked Amount			Recovery	=	100.34%	
50) Toluene-d8	10.09	98	1428996	46.69	ug/l	0.00
Spiked Amount			Recovery	=	93.38%	
62) 4-Bromofluorobenzene	12.40	95	380092	37.59	ug/l	0.00
Spiked Amount			Recovery	=	75.18%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	2947	Below Cal		93
4) Vinyl Chloride	2.18	62	9880	0.984 ug/l		98
5) Bromomethane	2.57	94	3357	Below Cal		92
6) Chloroethane	2.70	64	393	Below Cal	#	43
10) Methyl Iodide	3.95	142	1106	7.117 ug/l	#	53
11) Tert butyl alcohol	4.79	59	208	0.451 ug/l	#	72
14) Allyl chloride	4.07	41	4174	0.369 ug/l	#	63
16) Acetone	3.82	43	15804	6.078 ug/l		100
18) Methyl Acetate	4.34	43	5448	Below Cal	#	78
20) Methylene Chloride	4.55	84	2915	Below Cal		93
21) trans-1,2-Dichloroethene	5.06	96	4002	0.508 ug/l		92
25) 2-Butanone	6.85	43	2483	0.717 ug/l	#	1
27) cis-1,2-Dichloroethene	6.83	96	385639	43.929 ug/l		92
29) Tetrahydrofuran	7.21	42	518	0.286 ug/l	#	33
31) Cyclohexane	7.67	56	11993	0.595 ug/l	#	1
36) 1,1-Dichloropropene	7.67	75	45476	4.011 ug/l	#	49
38) Carbon Tetrachloride	7.67	117	59882	5.091 ug/l	#	17
43) Isopropyl Acetate	8.29	43	50497	4.323 ug/l	#	60
44) Trichloroethene	8.84	130	176116	19.061 ug/l		98
55) 1,1,2-Trichloroethane	10.63	97	8226	1.069 ug/l	#	1
60) Dibromochloromethane	10.63	129	1027718	121.028 ug/l	#	10
64) Tetrachloroethene	10.63	164	1166241	143.002 ug/l		100
75) 1,1,2,2-Tetrachloroethane	12.41	83	287	Below Cal	#	1
76) 1,2,3-Trichloropropane	12.40	75	179439	36.878 ug/l	#	100
81) trans-1,4-Dichloro-2-buten	12.40	75	179439	133.434 ug/l	#	13
95) Naphthalene	15.14	128	534	3.904 ug/l	#	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081418\
Data File : VN050613.D
Acq On : 14 Aug 2018 19:43
Operator : MD\SY
Sample : J4469-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)

Quant Time: Aug 15 07:11:13 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :


960-IW-21(17)

Tgt Ion:168 Resp: 670176

Ion Ratio Lower Upper




168 100

99 51.1 40.8 61.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#4

Vinyl Chloride

Concen: 0.984 ug/l

RT: 2.18 min Scan# 184

Delta R.T. 0.00 min

Lab File: VN050613.D


Acq: 14 Aug 2018 19:43

Tgt Ion: 62 Resp: 9880

Ion Ratio Lower Upper



62 100


64 32.6 25.2 37.8



Abundance Ion 61.90 (61.60 to 62.60): VNO

Ion 63.90 (63.60 to 64.60): VNO





#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 306

Delta R.T. 0.01 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :


960-IW-21(17)

Tgt Ion: 94 Resp: 3357

Ion Ratio Lower Upper




94 100

96 85.2 74.0 111.0



Abundance Ion 93.90 (93.60 to 94.60): VN0

Ion 95.90 (95.60 to 96.60): VN0



#6

Chloroethane

Concen: Below Cal

RT: 2.70 min Scan# 345

Delta R.T. 0.00 min

Lab File: VN050613.D


Acq: 14 Aug 2018 19:43

Tgt Ion: 64 Resp: 393

Ion Ratio Lower Upper



64 100


66 0.0 25.7 38.5#



Abundance Ion 63.90 (63.60 to 64.60): VN0

Ion 65.90 (65.60 to 66.60): VN0









#10
Methyl Iodide
Concen: 7.117 ug/l
RT: 3.95 min Scan# 734
Delta R.T. -0.00 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43




Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)


Tgt Ion: 142 Resp: 1106
Ion Ratio Lower Upper
142 100
127 9.0 32.6 49.0#
141 0.0 11.5 17.3#



#11
Tert butyl alcohol
Concen: 0.451 ug/l
RT: 4.79 min Scan# 994
Delta R.T. -0.01 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Tgt Ion: 59 Resp: 208
Ion Ratio Lower Upper
59 100
57 0.0 8.4 12.6#





#14

Allyl chloride

Concen: 0.369 ug/l

RT: 4.07 min Scan# 771

Delta R.T. -0.25 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :

960-IW-21(17)


Tgt Ion: 41 Resp: 4174

Ion Ratio Lower Upper

41 100

39 82.7 51.4 77.0#




76 0.0 29.4 44.0#



Abundance Ion 41.00 (40.70 to 41.70): VN0

Ion 39.00 (38.70 to 39.70): VN0

Ion 75.90 (75.60 to 76.60): VN0



#16

Acetone

Concen: 6.078 ug/l

RT: 3.82 min Scan# 695

Delta R.T. 0.01 min

Lab File: VN050613.D


Acq: 14 Aug 2018 19:43

Tgt Ion: 43 Resp: 15804

Ion Ratio Lower Upper



43 100


58 33.6 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0









#20
Methylene Chloride
Concen: Below Cal
RT: 4.55 min Scan# 921
Delta R.T. 0.00 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43




Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)


Tgt Ion:	84	Resp:	2915
Ion	Ratio	Lower	Upper
84	100		
49	119.8	92.6	138.8
51	34.2	28.6	43.0
86	52.2	52.2	78.2



#21
trans-1,2-Dichloroethene
Concen: 0.508 ug/l
RT: 5.06 min Scan# 1078
Delta R.T. 0.01 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Tgt Ion:	96	Resp:	4002
Ion	Ratio	Lower	Upper
96	100		
61	153.6	111.2	166.8
98	64.9	51.6	77.4





#25

2-Butanone

Concen: 0.717 ug/l

RT: 6.85 min Scan# 1637

Delta R.T. 0.01 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampled :


960-IW-21(17)

Tgt Ion: 43 Resp: 2483

Ion Ratio Lower Upper




43 100

72 90.9 21.8 32.6#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 72.00 (71.70 to 72.70): VN0



#27

cis-1,2-Dichloroethene

Concen: 43.929 ug/l

RT: 6.83 min Scan# 1631

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43


Tgt Ion: 96 Resp: 385639

Ion Ratio Lower Upper

96 100

61 124.4 0.0 278.2



98 64.6 0.0 128.8




Abundance Ion 95.90 (95.60 to 96.60): VN0

Ion 60.90 (60.60 to 61.60): VN0

Ion 97.90 (97.60 to 98.60): VN0





#29

Tetrahydrofuran

Concen: 0.286 ug/l

RT: 7.21 min Scan# 1749

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :

960-IW-21(17)





Tgt Ion: 42 Resp: 518

Ion Ratio Lower Upper

42 100

72 0.0 35.8 53.6#

71 0.0 33.4 50.0#



#31

Cyclohexane

Concen: 0.595 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.02 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43




Tgt Ion: 56 Resp: 11993


Ion Ratio Lower Upper

56 100

69 185.9 25.8 38.6#

84 156.2 67.8 101.6#









#33
 1,2-Dichloroethane-d4
 Concen: 50.334 ug/l
 RT: 8.03 min Scan# 2002
 Delta R.T. 0.00 min
 Lab File: VN050613.D
 Acq: 14 Aug 2018 19:43




Instrument :
 MSVOA_N
 ClientSampleId :
 960-IW-21(17)


Tgt Ion: 65 Resp: 425202
 Ion Ratio Lower Upper
 65 100
 67 54.2 0.0 109.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2176
 Delta R.T. 0.00 min
 Lab File: VN050613.D
 Acq: 14 Aug 2018 19:43

Tgt Ion: 114 Resp: 1018571
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 40.0
 88 15.4 0.0 30.8









#35
Dibromofluoromethane
Concen: 50.168 ug/l
RT: 7.59 min Scan# 1867
Delta R.T. 0.00 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43




Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)


Tgt Ion	Ratio	Lower	Upper
113	100		
111	100.2	81.0	121.6
192	22.2	17.6	26.4



#36
1,1-Dichloropropene
Concen: 4.011 ug/l
RT: 7.67 min Scan# 1891
Delta R.T. -0.12 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Tgt Ion	Ratio	Lower	Upper
75	100		
110	8.9	18.3	54.9#
77	0.0	25.0	37.4#









#38
Carbon Tetrachloride
Concen: 5.091 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.11 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)

Tgt Ion: 117 Resp: 59882
Ion Ratio Lower Upper
117 100
119 6.1 76.6 115.0#
121 0.0 25.0 37.6#




Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V






#43
Isopropyl Acetate
Concen: 4.323 ug/l
RT: 8.29 min Scan# 2083
Delta R.T. 0.12 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Tgt Ion: 43 Resp: 50497
Ion Ratio Lower Upper
43 100
61 0.0 16.2 24.2#
87 0.0 10.9 16.3#



Abundance Ion 43.00 (42.70 to 43.70): V
Ion 61.00 (60.70 to 61.70): V
Ion 87.00 (86.70 to 87.70): V





#44

Trichloroethene

Concen: 19.061 ug/l

RT: 8.84 min Scan# 2254

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :


960-IW-21(17)

Tgt Ion:130 Resp: 176116

Ion Ratio Lower Upper

130 100

95 95.1 0.0 193.8





Abundance Ion 129.80 (129.50 to 130.50): VN050587.D (-2237) (-)

Ion 94.90 (94.60 to 95.60): VN050587.D (-2237) (-)

8.84

8.75 8.80 8.85 8.90

Time-->



#50

Toluene-d8

Concen: 46.692 ug/l

RT: 10.09 min Scan# 2644

Delta R.T. 0.00 min

Lab File: VN050613.D


Acq: 14 Aug 2018 19:43

Tgt Ion: 98 Resp: 1428996

Ion Ratio Lower Upper

98 100

100 63.0 51.8 77.8




Abundance Ion 98.00 (97.70 to 98.70): VN050587.D (-2627) (-)


Ion 100.00 (99.70 to 100.70): VN050587.D (-2627) (-)

10.09

10.00 10.10 10.20 10.30

Time-->









#55
 1,1,2-Trichloroethane
 Concen: 1.069 ug/l
 RT: 10.63 min Scan# 2811
 Delta R.T. 0.06 min
 Lab File: VN050613.D
 Acq: 14 Aug 2018 19:43

Instrument :
 MSVOA_N
 ClientSampleId :
 960-IW-21(17)

Tgt Ion: 97 Resp: 8226
 Ion Ratio Lower Upper
 97 100
 83 275.6 68.5 102.7#
 85 35.6 44.6 66.8#
 99 18.3 49.1 73.7#




Abundance Ion 96.90 (96.60 to 97.60): VN0
 Ion 82.90 (82.60 to 83.60): VN0
 Ion 84.90 (84.60 to 85.60): VN0
 Ion 98.90 (98.60 to 99.60): VN0






#60
 Dibromochloromethane
 Concen: 121.028 ug/l
 RT: 10.63 min Scan# 2812
 Delta R.T. -0.27 min
 Lab File: VN050613.D
 Acq: 14 Aug 2018 19:43

Tgt Ion: 129 Resp: 1027718
 Ion Ratio Lower Upper
 129 100
 127 0.0 38.9 116.7#



Abundance Ion 128.80 (128.50 to 129.50): V
 Ion 126.80 (126.50 to 127.50): V









#62
4-Bromofluorobenzene
Concen: 37.593 ug/l
RT: 12.40 min Scan# 3363
Delta R.T. 0.00 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43




Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)


Tgt Ion: 95 Resp: 380092
Ion Ratio Lower Upper
95 100
174 91.2 0.0 177.8
176 88.1 0.0 175.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 3054
Delta R.T. 0.00 min
Lab File: VN050613.D
Acq: 14 Aug 2018 19:43

Tgt Ion: 117 Resp: 878169
Ion Ratio Lower Upper
117 100
82 53.0 42.4 63.6
119 32.2 25.8 38.8





#64

Tetrachloroethene

Concen: 143.002 ug/l

RT: 10.63 min Scan# 2812

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :

960-IW-21(17)

Tgt Ion:164 Resp: 1166241



Ion Ratio Lower Upper

164 100

166 127.7 102.1 153.1

129 91.4 72.7 109.1

131 87.9 69.9 104.9





Abundance

Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. 0.00 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43



Tgt Ion:152 Resp: 298973

Ion Ratio Lower Upper

152 100

115 56.4 28.1 84.2

150 156.7 0.0 347.8





Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.41 min Scan# 3364

Delta R.T. -0.10 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :

960-IW-21(17)


Tgt Ion: 83 Resp: 287

Ion Ratio Lower Upper

83 100

131 156.8 5.3 15.9#




85 0.0 32.1 96.5#



Abundance Ion 82.90 (82.60 to 83.60): VNO

Ion 130.80 (130.50 to 131.50): VNO

Ion 84.90 (84.60 to 85.60): VNO



#76

1,2,3-Trichloropropane

Concen: 36.878 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050613.D


Acq: 14 Aug 2018 19:43

Tgt Ion: 75 Resp: 179439

Ion Ratio Lower Upper



75 100


77 1.2 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VNO

Ion 77.00 (76.70 to 77.70): VNO





#81

trans-1,4-Dichloro-2-butene

Concen: 133.434 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43

Instrument :

MSVOA_N

ClientSampleId :

960-IW-21(17)


Tgt Ion: 75 Resp: 179439

Ion Ratio Lower Upper

75 100

53 0.2 72.2 108.2#




89 0.0 36.3 54.5#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#95

Naphthalene

Concen: 3.904 ug/l

RT: 15.14 min Scan# 4214

Delta R.T. 0.01 min

Lab File: VN050613.D

Acq: 14 Aug 2018 19:43


Tgt Ion: 128 Resp: 534

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#

129 0.0 8.5 12.7#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

