

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081418\
 Data File : VN050625.D
 Acq On : 15 Aug 2018 1:28
 Operator : MD\SY
 Sample : J4465-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 943-MW-04(23)

Quant Time: Aug 15 07:27:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	666812	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1043066	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	337596	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	437191	52.01	ug/l	0.00
Spiked Amount			Recovery	=	104.02%	
35) Dibromofluoromethane	7.59	113	418384	50.24	ug/l	0.00
Spiked Amount			Recovery	=	100.48%	
50) Toluene-d8	10.09	98	1494013	47.67	ug/l	0.00
Spiked Amount			Recovery	=	95.34%	
62) 4-Bromofluorobenzene	12.40	95	415878	40.17	ug/l	0.00
Spiked Amount			Recovery	=	80.34%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	2.06	50	3337	Below Cal		89
5) Bromomethane	2.57	94	3766	Below Cal		91
6) Chloroethane	2.69	64	246	Below Cal	#	43
10) Methyl Iodide	3.94	142	525	7.029	ug/l #	87
11) Tert butyl alcohol	4.78	59	257	0.560	ug/l #	72
16) Acetone	3.83	43	13696	5.106	ug/l #	87
18) Methyl Acetate	4.33	43	4108	Below Cal	#	69
20) Methylene Chloride	4.54	84	1967	Below Cal		90
25) 2-Butanone	6.86	43	2369	0.687	ug/l #	83
27) cis-1,2-Dichloroethene	6.83	96	1779	0.204	ug/l	68
29) Tetrahydrofuran	7.22	42	579	0.322	ug/l #	37
31) Cyclohexane	7.67	56	13855	0.747	ug/l #	1
36) 1,1-Dichloropropene	7.67	75	44512	3.834	ug/l #	49
38) Carbon Tetrachloride	7.67	117	58711	4.874	ug/l #	17
43) Isopropyl Acetate	8.25	43	12604	1.054	ug/l #	60
44) Trichloroethene	8.84	130	2764	0.292	ug/l	80
48) Methyl methacrylate	9.43	41	2029	0.337	ug/l #	22
60) Dibromochloromethane	10.63	129	27521	3.165	ug/l #	10
64) Tetrachloroethene	10.63	164	31047	3.603	ug/l	92
71) Bromoform	12.41	173	3642	0.635	ug/l #	1
75) 1,1,2,2-Tetrachloroethane	12.50	83	129	Below Cal	#	26
76) 1,2,3-Trichloropropane	12.40	75	196273	35.690	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.40	75	196273	129.254	ug/l #	13
95) Naphthalene	15.14	128	795	3.917	ug/l #	69


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\TERASTORAGE\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081418\
Data File : VN050625.D
Acq On : 15 Aug 2018 1:28
Operator : MD\SY
Sample : J4465-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
943-MW-04(23)

Quant Time: Aug 15 07:27:25 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1891

Delta R.T. 0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion:168 Resp: 666812

Ion Ratio Lower Upper

168 100

99 50.4 40.8 61.2



Abundance Ion 167.90 (167.60 to 168.60): VN050587.D (-1871) (-)

Ion 98.90 (98.60 to 99.60): VN050587.D (-1871) (-)

7.67

250000

200000

150000



100000

50000

0

7.60 7.70 7.80

Time-->



#5

Bromomethane

Concen: Below Cal

RT: 2.57 min Scan# 306

Delta R.T. 0.01 min

Lab File: VN050625.D


Acq: 15 Aug 2018 1:28

Tgt Ion: 94 Resp: 3766

Ion Ratio Lower Upper

94 100

96 83.5 74.0 111.0



Abundance Ion 93.90 (93.60 to 94.60): VN050587.D (-287) (-)

Ion 95.90 (95.60 to 96.60): VN050587.D (-287) (-)

2.57

1500


1000


500

0

2.50 2.55 2.60 2.65

Time-->





#6

Chloroethane

Concen: Below Cal

RT: 2.69 min Scan# 341

Delta R.T. -0.01 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion: 64 Resp: 246

Ion Ratio Lower Upper

64 100

66 0.0 25.7 38.5#






Abundance Ion 63.90 (63.60 to 64.60): VN050587.D (-329) (-)

Ion 65.90 (65.60 to 66.60): VN050587.D (-329) (-)

2.69

2.68

2.70



#10

Methyl Iodide

Concen: 7.029 ug/l

RT: 3.94 min Scan# 732

Delta R.T. -0.01 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 142 Resp: 525

Ion Ratio Lower Upper

142 100

127 37.5 32.6 49.0

141 0.0 11.5 17.3#



Abundance Ion 141.80 (141.50 to 142.50): VN050587.D (-713) (-)



Ion 126.80 (126.50 to 127.50): VN050587.D (-713) (-)


Ion 140.80 (140.50 to 141.50): VN050587.D (-713) (-)

3.94

3.92

3.96





#11

Tert butyl alcohol

Concen: 0.560 ug/l

RT: 4.78 min Scan# 993

Delta R.T. -0.01 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :





943-MW-04(23)

Tgt Ion: 59 Resp: 257

Ion Ratio Lower Upper

59 100

57 0.0 8.4 12.6#



#16

Acetone

Concen: 5.106 ug/l

RT: 3.83 min Scan# 696

Delta R.T. 0.01 min

Lab File: VN050625.D




Acq: 15 Aug 2018 1:28


Tgt Ion: 43 Resp: 13696

Ion Ratio Lower Upper

43 100

58 26.4 27.1 40.7#









#18
Methyl Acetate
Concen: Below Cal
RT: 4.33 min Scan# 853
Delta R.T. 0.01 min
Lab File: VN050625.D
Acq: 15 Aug 2018 1:28




Instrument :
MSVOA_N
ClientSampleId :
943-MW-04(23)


Tgt Ion: 43 Resp: 4108
Ion Ratio Lower Upper
43 100
74 9.2 19.7 29.5#



#20
Methylene Chloride
Concen: Below Cal
RT: 4.54 min Scan# 919
Delta R.T. -0.01 min
Lab File: VN050625.D
Acq: 15 Aug 2018 1:28

Tgt Ion: 84 Resp: 1967
Ion Ratio Lower Upper
84 100
49 95.3 92.6 138.8
51 35.1 28.6 43.0
86 65.1 52.2 78.2









#25
 2-Butanone
 Concen: 0.687 ug/l
 RT: 6.86 min Scan# 1638
 Delta R.T. 0.02 min
 Lab File: VN050625.D
 Acq: 15 Aug 2018 1:28




Instrument :
 MSVOA_N
 ClientSampleId :
 943-MW-04(23)


Tgt Ion: 43 Resp: 2369
 Ion Ratio Lower Upper
 43 100
 72 18.1 21.8 32.6#



#27
 cis-1,2-Dichloroethene
 Concen: 0.204 ug/l
 RT: 6.83 min Scan# 1630
 Delta R.T. -0.00 min
 Lab File: VN050625.D
 Acq: 15 Aug 2018 1:28

Tgt Ion: 96 Resp: 1779
 Ion Ratio Lower Upper
 96 100
 61 192.2 0.0 278.2
 98 69.1 0.0 128.8





#29

Tetrahydrofuran

Concen: 0.322 ug/l

RT: 7.22 min Scan# 1751

Delta R.T. 0.01 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :

943-MW-04(23)


Tgt Ion: 42 Resp: 579

Ion Ratio Lower Upper

42 100

72 5.4 35.8 53.6#

71 0.0 33.4 50.0#



Abundance Ion 42.00 (41.70 to 42.70): VN0

Ion 72.00 (71.70 to 72.70): VN0

Ion 71.00 (70.70 to 71.70): VN0

7.22

400



300

200

100

0

Time-->



#31

Cyclohexane

Concen: 0.747 ug/l

RT: 7.67 min Scan# 1890

Delta R.T. 0.01 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 56 Resp: 13855

Ion Ratio Lower Upper

56 100

69 167.9 25.8 38.6#

84 130.6 67.8 101.6#



Abundance Ion 56.00 (55.70 to 56.70): VN0

Ion 69.00 (68.70 to 69.70): VN0

Ion 84.00 (83.70 to 84.70): VN0


7.67


10000

5000

0

Time-->





#33

1,2-Dichloroethane-d4

Concen: 52.014 ug/l

RT: 8.03 min Scan# 2003

Delta R.T. 0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion: 65 Resp: 437191

Ion Ratio Lower Upper

65 100

67 54.2 0.0 109.8



Abundance Ion 64.90 (64.60 to 65.60): VN050587.D (-1983) (-)

Ion 66.90 (66.60 to 67.60): VN050587.D (-1983) (-)

8.03

200000



150000

100000

50000

0

Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2176

Delta R.T. -0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 114 Resp: 1043066

Ion Ratio Lower Upper

114 100

63 19.4 0.0 40.0

88 14.9 0.0 30.8



Abundance Ion 113.90 (113.60 to 114.60): VN050587.D (-2157) (-)

Ion 63.00 (62.70 to 63.70): VN050587.D (-2157) (-)

Ion 87.90 (87.60 to 88.60): VN050587.D (-2157) (-)

8.59


600000


400000

200000

0

Time-->









#35
 Dibromofluoromethane
 Concen: 50.241 ug/l
 RT: 7.59 min Scan# 1867
 Delta R.T. -0.00 min
 Lab File: VN050625.D
 Acq: 15 Aug 2018 1:28




Instrument :
 MSVOA_N
 ClientSampleId :
 943-MW-04(23)


Tgt Ion:113 Resp: 418384
 Ion Ratio Lower Upper
 113 100
 111 101.6 81.0 121.6
 192 22.2 17.6 26.4



#36
 1,1-Dichloropropene
 Concen: 3.834 ug/l
 RT: 7.67 min Scan# 1890
 Delta R.T. -0.13 min
 Lab File: VN050625.D
 Acq: 15 Aug 2018 1:28

Tgt Ion: 75 Resp: 44512
 Ion Ratio Lower Upper
 75 100
 110 8.9 18.3 54.9#
 77 0.0 25.0 37.4#









#38
Carbon Tetrachloride
Concen: 4.874 ug/l
RT: 7.67 min Scan# 1890
Delta R.T. -0.11 min
Lab File: VN050625.D
Acq: 15 Aug 2018 1:28




Instrument :
MSVOA_N
ClientSampleId :
943-MW-04(23)


Tgt Ion: 117 Resp: 58711
Ion Ratio Lower Upper
117 100
119 6.5 76.6 115.0#
121 0.0 25.0 37.6#



#43
Isopropyl Acetate
Concen: 1.054 ug/l
RT: 8.25 min Scan# 2070
Delta R.T. 0.08 min
Lab File: VN050625.D
Acq: 15 Aug 2018 1:28

Tgt Ion: 43 Resp: 12604
Ion Ratio Lower Upper
43 100
61 0.0 16.2 24.2#
87 0.0 10.9 16.3#





#44

Trichloroethene

Concen: 0.292 ug/l

RT: 8.84 min Scan# 2255

Delta R.T. 0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion:130 Resp: 2764




Ion Ratio Lower Upper

130 100

95 77.0 0.0 193.8



Abundance Ion 129.80 (129.50 to 130.50): VN050625.D (-2161) (-)



#48

Methyl methacrylate

Concen: 0.337 ug/l

RT: 9.43 min Scan# 2439

Delta R.T. 0.23 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 41 Resp: 2029

Ion Ratio Lower Upper



41 100


69 0.0 73.4 110.0#

39 19.7 43.0 64.6#



Abundance Ion 41.00 (40.70 to 41.70): VN050625.D (-2274) (-)





#50

Toluene-d8

Concen: 47.670 ug/l

RT: 10.09 min Scan# 2644

Delta R.T. -0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion: 98 Resp: 1494013

Ion Ratio Lower Upper

98 100

100 63.4 51.8 77.8



Abundance Ion 98.00 (97.70 to 98.70): VN050587.D (-2627) (-)

Ion 100.00 (99.70 to 100.70): VN050587.D (-2627) (-)

10.09

800000



600000

400000

200000

0

Time-->



#60

Dibromochloromethane

Concen: 3.165 ug/l

RT: 10.63 min Scan# 2812

Delta R.T. -0.27 min

Lab File: VN050625.D


Acq: 15 Aug 2018 1:28

Tgt Ion: 129 Resp: 27521

Ion Ratio Lower Upper

129 100

127 0.0 38.9 116.7#



Abundance Ion 128.80 (128.50 to 129.50): VN050587.D (-2881) (-)

Ion 126.80 (126.50 to 127.50): VN050587.D (-2881) (-)

10.63


15000


10000

5000

0

Time-->





#62

4-Bromofluorobenzene

Concen: 40.167 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :

943-MW-04(23)


Tgt Ion: 95 Resp: 415878

Ion Ratio Lower Upper

95 100

174 91.3 0.0 177.8




176 88.5 0.0 175.0



Abundance Ion 94.90 (94.60 to 95.60): VN0

Ion 173.80 (173.50 to 174.50): VN0

Ion 175.80 (175.50 to 176.50): VN0



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 3054

Delta R.T. -0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 117 Resp: 927753

Ion Ratio Lower Upper

117 100

82 53.3 42.4 63.6



119 32.4 25.8 38.8




Abundance Ion 116.90 (116.60 to 117.60): VN0

Ion 82.00 (81.70 to 82.70): VN0

Ion 118.90 (118.60 to 119.60): VN0





#64

Tetrachloroethene

Concen: 3.603 ug/l

RT: 10.63 min Scan# 2812

Delta R.T. -0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :

943-MW-04(23)

Tgt Ion:164 Resp: 31047


Ion Ratio Lower Upper

164 100

166 135.6 102.1 153.1

129 95.1 72.7 109.1

131 99.1 69.9 104.9






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#71

Bromoform

Concen: 0.635 ug/l

RT: 12.41 min Scan# 3364

Delta R.T. 0.28 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion:173 Resp: 3642

Ion Ratio Lower Upper

173 100

175 647.2 24.4 73.2#



254 0.0 0.0 0.0




Abundance Ion 172.70 (172.40 to 173.40): V

Ion 174.70 (174.40 to 175.40): V

Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3656

Delta R.T. -0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :

943-MW-04(23)


Tgt Ion:152 Resp: 337596

Ion Ratio Lower Upper

152 100

115 57.2 28.1 84.2




150 156.8 0.0 347.8



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#75

1,1,2,2-Tetrachloroethane

Concen: Below Cal

RT: 12.50 min Scan# 3394

Delta R.T. 0.00 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 83 Resp: 129

Ion Ratio Lower Upper

83 100

131 0.0 5.3 15.9#



85 0.0 32.1 96.5#




Abundance Ion 82.90 (82.60 to 83.60): V

Ion 130.80 (130.50 to 131.50): V

Ion 84.90 (84.60 to 85.60): V





#76

1,2,3-Trichloropropane

Concen: 35.690 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. -0.15 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :


943-MW-04(23)

Tgt Ion: 75 Resp: 196273

Ion Ratio Lower Upper




75 100

77 1.5 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#81

trans-1,4-Dichloro-2-butene

Concen: 129.254 ug/l

RT: 12.40 min Scan# 3363

Delta R.T. 0.10 min

Lab File: VN050625.D

Acq: 15 Aug 2018 1:28


Tgt Ion: 75 Resp: 196273

Ion Ratio Lower Upper

75 100

53 0.1 72.2 108.2#



89 0.0 36.3 54.5#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0





#95

Naphthalene

Concen: 3.917 ug/l

RT: 15.14 min Scan# 4213

Delta R.T. 0.00 min

Lab File: VN050625.D


Acq: 15 Aug 2018 1:28

Instrument :

MSVOA_N

ClientSampleId :

943-MW-04(23)




Tgt Ion:128 Resp: 795

Ion Ratio Lower Upper

128 100

127 0.0 10.3 15.5#

129 0.0 8.5 12.7#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

