

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050601.D
 Acq On : 14 Aug 2018 14:46
 Operator : MD\SY
 Sample : PB112007TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112007TB

Quant Time: Aug 15 09:05:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	653433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1052720	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	927852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	338180	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	436416	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
35) Dibromofluoromethane	7.59	113	411650	48.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.96%	
50) Toluene-d8	10.09	98	1465240	46.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.64%	
62) 4-Bromofluorobenzene	12.40	95	414947	39.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.83	43	12868	4.84	ug/l	92
20) Methylene Chloride	4.55	84	12696	0.36	ug/l	96
43) Isopropyl Acetate	8.17	43	202272	16.76	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050601.D
Acq On : 14 Aug 2018 14:46
Operator : MD\SY
Sample : PB112007TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112007TB

Quant Time: Aug 15 09:05:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

