

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050603.D
 Acq On : 14 Aug 2018 15:36
 Operator : MD\SY
 Sample : PB112004TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB112004TB

Quant Time: Aug 15 11:30:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	597481	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1007502	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	870412	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	324065	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	415374	55.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.30%	
35) Dibromofluoromethane	7.59	113	387084	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	10.09	98	1367895	45.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.38%	
62) 4-Bromofluorobenzene	12.40	95	389517	38.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.90%	
Target Compounds						
16) Acetone	3.82	43	13128	5.56	ug/l	91
20) Methylene Chloride	4.55	84	13666	0.65	ug/l	92
43) Isopropyl Acetate	8.17	43	225572	19.52	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050603.D
Acq On : 14 Aug 2018 15:36
Operator : MD\SY
Sample : PB112004TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB112004TB

Quant Time: Aug 15 11:30:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

