

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050607.D
 Acq On : 14 Aug 2018 17:14
 Operator : MD\SY
 Sample : J4442-50
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-1-5

Quant Time: Aug 15 12:33:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	670985	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1094140	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	940191	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	347158	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	459852	54.37	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	108.74%	
35) Dibromofluoromethane	7.59	113	429995	49.23	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	98.46%	
50) Toluene-d8	10.09	98	1511279	45.97	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	91.94%	
62) 4-Bromofluorobenzene	12.40	95	418445	38.53	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	77.06%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	17132	6.70	ug/l	100
20) Methylene Chloride	4.55	84	34217	3.06	ug/l	99
30) Chloroform	7.37	83	28297	1.86	ug/l	98
43) Isopropyl Acetate	8.17	43	236754	18.87	ug/l	98
47) Bromodichloromethane	9.40	83	6885	0.55	ug/l	97
95) Naphthalene	15.13	128	39127	6.56	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050607.D
Acq On : 14 Aug 2018 17:14
Operator : MD\SY
Sample : J4442-50
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-1-5

Quant Time: Aug 15 12:33:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

