

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050609.D
 Acq On : 14 Aug 2018 18:04
 Operator : MD\SY
 Sample : J4469-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 956-IW-13(19)

Quant Time: Aug 15 12:40:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	688391	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1038805	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	917106	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	352173	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	431307	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
35) Dibromofluoromethane	7.59	113	403724	48.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.36%	
50) Toluene-d8	10.09	98	1441603	46.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.38%	
62) 4-Bromofluorobenzene	12.40	95	411140	39.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	79.74%	
Target Compounds						
4) Vinyl Chloride	2.18	62	6001	0.58	ug/l	96
16) Acetone	3.82	43	16869	6.37	ug/l	96
21) trans-1,2-Dichloroethene	5.05	96	14756	1.82	ug/l	90
27) cis-1,2-Dichloroethene	6.83	96	1139205	126.34	ug/l	92
44) Trichloroethene	8.84	130	2131186	226.16	ug/l	100
64) Tetrachloroethene	10.63	164	10697130	1255.97	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050609.D
Acq On : 14 Aug 2018 18:04
Operator : MD\SY
Sample : J4469-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
956-IW-13(19)

Quant Time: Aug 15 12:40:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

