

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050613.D
 Acq On : 14 Aug 2018 19:43
 Operator : MD\SY
 Sample : J4469-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 960-IW-21(17)

Quant Time: Aug 15 13:32:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	670176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1018571	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	878169	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	298973	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	425202	50.33	ug/l	0.00
Spiked Amount			Recovery	=	100.66%	
35) Dibromofluoromethane	7.59	113	407961	50.17	ug/l	0.00
Spiked Amount			Recovery	=	100.34%	
50) Toluene-d8	10.09	98	1428996	46.69	ug/l	0.00
Spiked Amount			Recovery	=	93.38%	
62) 4-Bromofluorobenzene	12.40	95	380092	37.59	ug/l	0.00
Spiked Amount			Recovery	=	75.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.18	62	9880	0.98	ug/l	98
16) Acetone	3.82	43	15804	6.08	ug/l	100
27) cis-1,2-Dichloroethene	6.83	96	385639	43.93	ug/l	92
44) Trichloroethene	8.84	130	176116	19.06	ug/l	98
64) Tetrachloroethene	10.63	164	1166241	143.00	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050613.D
Acq On : 14 Aug 2018 19:43
Operator : MD\SY
Sample : J4469-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
960-IW-21(17)

Quant Time: Aug 15 13:32:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

