

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050615.D
 Acq On : 14 Aug 2018 20:32
 Operator : MD\SY
 Sample : J4465-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 944-MW-05(17)

Quant Time: Aug 15 13:37:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	680263	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1090966	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	944079	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	344865	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	457048	53.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.60%	
35) Dibromofluoromethane	7.59	113	426082	48.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.84%	
50) Toluene-d8	10.09	98	1534498	46.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.62%	
62) 4-Bromofluorobenzene	12.40	95	419345	38.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	8599	2.58	ug/l	91
44) Trichloroethene	8.84	130	7105	0.72	ug/l	94
64) Tetrachloroethene	10.63	164	369866	42.19	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050615.D
Acq On : 14 Aug 2018 20:32
Operator : MD\SY
Sample : J4465-05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
944-MW-05(17)

Quant Time: Aug 15 13:37:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

