

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050621.D
 Acq On : 14 Aug 2018 23:49
 Operator : MD\SY
 Sample : J4465-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 954-FB081318

Quant Time: Aug 15 14:55:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	609611	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	938062	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	795716	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	295511	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	396857	51.65	ug/l	0.00
Spiked Amount			Recovery	=	103.30%	
35) Dibromofluoromethane	7.59	113	373383	49.86	ug/l	0.00
Spiked Amount			Recovery	=	99.72%	
50) Toluene-d8	10.09	98	1330613	47.21	ug/l	0.00
Spiked Amount			Recovery	=	94.42%	
62) 4-Bromofluorobenzene	12.40	95	362354	38.91	ug/l	0.00
Spiked Amount			Recovery	=	77.82%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	30381	14.47	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050621.D
Acq On : 14 Aug 2018 23:49
Operator : MD\SY
Sample : J4465-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
954-FB081318

Quant Time: Aug 15 14:55:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

