

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050622.D
 Acq On : 15 Aug 2018 00:14
 Operator : MD\SY
 Sample : J4465-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 940-MW-01(23)

Manual Integrations
 APPROVED

MMDadoda
 8/15/2018 3:32:30 PM

Quant Time: Aug 15 15:01:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	688915	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1018194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	952170	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	369747	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	420725	48.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.90%	
35) Dibromofluoromethane	7.59	113	399591	49.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.32%	
50) Toluene-d8	10.09	98	1451638	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
62) 4-Bromofluorobenzene	12.40	95	436913	43.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.46%	
Target Compounds						
4) Vinyl Chloride	2.19	62	5355	0.52	ug/l	100
16) Acetone	3.82	43	14454	5.25	ug/l	95
27) cis-1,2-Dichloroethene	6.83	96	1242084	137.64	ug/l	92
44) Trichloroethene	8.84	130	1510898	163.58	ug/l	100
64) Tetrachloroethene	10.62	164	23778190m	2689.03	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050622.D
Acq On : 15 Aug 2018 00:14
Operator : MD\SY
Sample : J4465-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
940-MW-01(23)

Manual Integrations
APPROVED

MMDadoda
8/15/2018 3:32:30 PM

Quant Time: Aug 15 15:01:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

