

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050623.D
 Acq On : 15 Aug 2018 00:39
 Operator : MD\SY
 Sample : J4465-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 941-MW-02(23.8)

Manual Integrations
 APPROVED

MMDadoda
 8/15/2018 3:32:34 PM

Quant Time: Aug 15 14:19:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	678315	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1037731	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	905261	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	316149	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	435777	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
35) Dibromofluoromethane	7.59	113	412729	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
50) Toluene-d8	10.09	98	1463423	46.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.86%	
62) 4-Bromofluorobenzene	12.40	95	391458	38.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	76.00%	
Target Compounds						
4) Vinyl Chloride	2.18	62	9927	0.98	ug/l	98
16) Acetone	3.82	43	15860	6.01	ug/l	96
27) cis-1,2-Dichloroethene	6.83	96	266488	29.99	ug/l	90
44) Trichloroethene	8.84	130	325920	34.62	ug/l	99
64) Tetrachloroethene	10.63	164	727731	86.56	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050623.D
Acq On : 15 Aug 2018 00:39
Operator : MD\SY
Sample : J4465-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
941-MW-02(23.8)

Manual Integrations
APPROVED

MMDadoda
8/15/2018 3:32:34 PM

Quant Time: Aug 15 14:19:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

