

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050624.D
 Acq On : 15 Aug 2018 1:03
 Operator : MD\SY
 Sample : J4465-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 942-MW-03A(17)

Quant Time: Aug 15 14:21:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	675903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1012810	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	899145	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	324654	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	421891	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
35) Dibromofluoromethane	7.59	113	399230	49.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.74%	
50) Toluene-d8	10.09	98	1432729	47.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.16%	
62) 4-Bromofluorobenzene	12.40	95	394896	39.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	2.18	62	5085	0.50	ug/l	97
16) Acetone	3.82	43	11193	3.83	ug/l	91
27) cis-1,2-Dichloroethene	6.83	96	625651	70.67	ug/l	92
44) Trichloroethene	8.84	130	702210	76.43	ug/l	99
64) Tetrachloroethene	10.63	164	5723272	685.40	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050624.D
Acq On : 15 Aug 2018 1:03
Operator : MD\SY
Sample : J4465-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
942-MW-03A(17)

Quant Time: Aug 15 14:21:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

