

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050627.D
 Acq On : 15 Aug 2018 2:17
 Operator : MD\SY
 Sample : J4465-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 946-MW-07(22.5)

Quant Time: Aug 15 14:25:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	627482	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	989132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	855740	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	296229	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	421658	53.31	ug/l	0.00
Spiked Amount						
			Recovery	=	106.62%	
35) Dibromofluoromethane	7.59	113	398877	50.51	ug/l	0.00
Spiked Amount						
			Recovery	=	101.02%	
50) Toluene-d8	10.09	98	1406346	47.32	ug/l	0.00
Spiked Amount						
			Recovery	=	94.64%	
62) 4-Bromofluorobenzene	12.40	95	372657	37.95	ug/l	0.00
Spiked Amount						
			Recovery	=	75.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	8938	3.09	ug/l	98
64) Tetrachloroethene	10.63	164	75458	9.50	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050627.D
Acq On : 15 Aug 2018 2:17
Operator : MD\SY
Sample : J4465-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
946-MW-07(22.5)

Quant Time: Aug 15 14:25:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

