

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081418\
 Data File : VN050629.D
 Acq On : 15 Aug 2018 3:07
 Operator : MD\SY
 Sample : J4465-11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 948-MW-15(17)

Quant Time: Aug 15 14:28:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 08:07:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	592225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	943667	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	844926	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	299755	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	401647	53.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.60%	
35) Dibromofluoromethane	7.59	113	383120	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
50) Toluene-d8	10.09	98	1347020	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
62) 4-Bromofluorobenzene	12.40	95	370705	39.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.16%	
Target Compounds						
16) Acetone	3.82	43	12229	5.14	ug/l	# 81
64) Tetrachloroethene	10.63	164	50003	6.37	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081418\
Data File : VN050629.D
Acq On : 15 Aug 2018 3:07
Operator : MD\SY
Sample : J4465-11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
948-MW-15(17)

Quant Time: Aug 15 14:28:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 08:07:08 2018
Response via : Initial Calibration

