

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081419\
 Data File : VN057282.D
 Acq On : 14 Aug 2019 12:55
 Operator : JC/SP
 Sample : K4096-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-081219

Quant Time: Aug 14 13:51:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 11:11:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1070837	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1495224	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1220767	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	296412	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	513473	47.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.06%	
35) Dibromofluoromethane	7.58	113	453637	47.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.02%	
50) Toluene-d8	10.08	98	1741891	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
62) 4-Bromofluorobenzene	12.40	95	437144	36.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	73.22%	
Target Compounds						
16) Acetone	3.81	43	31914	8.617	ug/l	96
18) Methyl Acetate	4.32	43	26451	1.645	ug/l #	67
20) Methylene Chloride	4.55	84	45170	4.175	ug/l	96
43) Isopropyl Acetate	8.16	43	217004	12.541	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081419\
Data File : VN057282.D
Acq On : 14 Aug 2019 12:55
Operator : JC/SP
Sample : K4096-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-081219

Quant Time: Aug 14 13:51:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 11:11:44 2019
Response via : Initial Calibration

