

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081419\
 Data File : VN057283.D
 Acq On : 14 Aug 2019 13:20
 Operator : JC/SP
 Sample : K4089-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-03-081219

Quant Time: Aug 14 14:39:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 11:11:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1127208	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1611569	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	1340112	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	336723	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	544385	47.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.74%	
35) Dibromofluoromethane	7.58	113	483050	46.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.88%	
50) Toluene-d8	10.09	98	1863286	49.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.28%	
62) 4-Bromofluorobenzene	12.40	95	489820	38.06	ug/l	0.00
Spiked Amount	50.000		Recovery	=	76.12%	
Target Compounds						
16) Acetone	3.82	43	31088	7.974	ug/l	99
18) Methyl Acetate	4.33	43	26163	1.474	ug/l	94
20) Methylene Chloride	4.55	84	46494	4.083	ug/l	98
43) Isopropyl Acetate	8.16	43	174174	9.339	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081419\
Data File : VN057283.D
Acq On : 14 Aug 2019 13:20
Operator : JC/SP
Sample : K4089-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-03-081219

Quant Time: Aug 14 14:39:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 11:11:44 2019
Response via : Initial Calibration

