

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081419\
 Data File : VN057285.D
 Acq On : 14 Aug 2019 14:08
 Operator : JC/SP
 Sample : K4231-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 REACTOR-3-B-(2)

Quant Time: Aug 14 14:34:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 14 11:11:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	1097919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1564063	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1275111	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	311856	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	532170	48.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.10%	
35) Dibromofluoromethane	7.58	113	469681	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
50) Toluene-d8	10.09	98	1818694	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	12.40	95	467270	37.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	74.82%	
Target Compounds						
16) Acetone	3.81	43	55998	14.747	ug/l	95
18) Methyl Acetate	4.32	43	28092	1.747	ug/l	100
20) Methylene Chloride	4.55	84	68529	6.178	ug/l	97
43) Isopropyl Acetate	8.16	43	211905	11.707	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081419\
Data File : VN057285.D
Acq On : 14 Aug 2019 14:08
Operator : JC/SP
Sample : K4231-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
REACTOR-3-B-(2)

Quant Time: Aug 14 14:34:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 14 11:11:44 2019
Response via : Initial Calibration

