


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081419\
 Data File : VN057275.D
 Acq On : 14 Aug 2019 8:21
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N081319W.M
 Title : SW846 8260
 Last Update : Wed Aug 14 06:25:03 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3289

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.3	14193	PASS
75	95	30	60	47.6	44184	PASS
95	95	100	100	100.0	92917	PASS
96	95	5	9	6.3	5850	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	93.2	86616	PASS
175	174	5	9	7.9	6839	PASS
176	174	95	101	96.6	83645	PASS
177	176	5	9	6.4	5349	PASS