

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN081420\
 Data File : VN062957.D
 Acq On : 15 Aug 2020 00:59
 Operator : JC/MD
 Sample : L3682-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200514-2801-B

Quant Time: Aug 17 09:33:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	174363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	328978	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	355466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125183	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	150363	52.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.30%	
35) Dibromofluoromethane	7.55	113	110804	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	
50) Toluene-d8	10.06	98	443468	50.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.38	95	165101	49.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.46%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.68	59	1833	3.879	ug/l #	72
16) Acetone	3.78	43	48699	46.741	ug/l	98
17) Carbon Disulfide	4.00	76	7935	0.972	ug/l	100
18) Methyl Acetate	4.29	43	7710	2.045	ug/l #	65
20) Methylene Chloride	4.50	84	15443	7.452	ug/l	96
25) 2-Butanone	6.80	43	11485	7.783	ug/l	96
29) Tetrahydrofuran	7.16	42	267	0.279	ug/l #	42
31) Cyclohexane	7.63	56	3753	1.074	ug/l #	1
36) 1,1-Dichloropropene	7.63	75	14686	4.585	ug/l #	49
37) Ethyl Acetate	6.80	43	11485	3.432	ug/l #	78
38) Carbon Tetrachloride	7.63	117	17833	4.707	ug/l #	16
41) Methacrylonitrile	7.16	41	362	0.240	ug/l #	6
43) Isopropyl Acetate	8.13	43	28583	5.121	ug/l	96
48) Methyl methacrylate	9.15	41	4374	1.593	ug/l #	24
51) 4-Methyl-2-Pentanone	9.95	43	8735	2.419	ug/l	91
54) cis-1,3-Dichloropropene	9.88	75	17911	4.260	ug/l #	68
56) Ethyl methacrylate	10.52	69	1860	0.490	ug/l #	1
59) 2-Hexanone	10.74	43	11644	4.323	ug/l #	57
71) Bromoform	12.38	173	1276	0.523	ug/l #	1
74) N-amyl acetate	12.00	43	1628	0.426	ug/l #	51
76) 1,2,3-Trichloropropane	12.38	75	83534	24.743	ug/l #	100
81) trans-1,4-Dichloro-2-buten	12.38	75	83534	75.534	ug/l #	10
92) 1,2-Dibromo-3-Chloropropan	14.29	75	285	0.400	ug/l #	10
95) Naphthalene	15.12	128	29	3.713	ug/l #	68


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN081420\
Data File : VN062957.D
Acq On : 15 Aug 2020 00:59
Operator : JC/MD
Sample : L3682-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200514-2801-B

Quant Time: Aug 17 09:33:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :


20200514-2801-B

Tgt Ion:168 Resp: 174363

Ion Ratio Lower Upper

168 100

99 60.9 52.9 79.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO

7.63

80000



60000

40000

20000

0

Time--> 7.50 7.60 7.70



#11

Tert butyl alcohol

Concen: 3.879 ug/l

RT: 4.68 min Scan# 905

Delta R.T. -0.06 min

Lab File: VN062957.D


Acq: 15 Aug 2020 00:59

Tgt Ion: 59 Resp: 1833

Ion Ratio Lower Upper

59 100

57 0.0 8.3 12.5#



Abundance Ion 59.00 (58.70 to 59.70): VNO

Ion 57.00 (56.70 to 57.70): VNO

4.68


1500


1000

500

0

Time--> 4.62 4.64 4.66 4.68 4.70





#16

Acetone

Concen: 46.741 ug/l

RT: 3.78 min Scan# 624

Delta R.T. 0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :





20200514-2801-B

Tgt Ion: 43 Resp: 48699

Ion Ratio Lower Upper

43 100

58 31.2 24.1 36.1



#17

Carbon Disulfide

Concen: 0.972 ug/l

RT: 4.00 min Scan# 694

Delta R.T. 0.00 min

Lab File: VN062957.D




Acq: 15 Aug 2020 00:59


Tgt Ion: 76 Resp: 7935

Ion Ratio Lower Upper

76 100

78 9.5 7.7 11.5









#18
Methyl Acetate
Concen: 2.045 ug/l
RT: 4.29 min Scan# 784
Delta R.T. 0.02 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59




Instrument :
MSVOA_N
ClientSampled :
20200514-2801-B


Tgt Ion: 43 Resp: 7710
Ion Ratio Lower Upper
43 100
74 6.1 18.6 28.0#



#20
Methylene Chloride
Concen: 7.452 ug/l
RT: 4.50 min Scan# 848
Delta R.T. 0.00 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Tgt Ion: 84 Resp: 15443
Ion Ratio Lower Upper
84 100
49 122.6 103.4 155.2
51 38.3 29.3 43.9
86 63.9 49.9 74.9





#25

2-Butanone

Concen: 7.783 ug/l

RT: 6.80 min Scan# 1563

Delta R.T. 0.01 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :





20200514-2801-B

Tgt Ion: 43 Resp: 11485

Ion Ratio Lower Upper

43 100

72 24.2 21.1 31.7



#29

Tetrahydrofuran

Concen: 0.279 ug/l

RT: 7.16 min Scan# 1676

Delta R.T. -0.01 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59




Tgt Ion: 42 Resp: 267


Ion Ratio Lower Upper

42 100

72 11.2 35.0 52.4#

71 0.0 33.2 49.8#









#31
Cyclohexane
Concen: 1.074 ug/l
RT: 7.63 min Scan# 1822
Delta R.T. 0.02 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Instrument :
MSVOA_N
ClientSampleId :
20200514-2801-B

Tgt Ion: 56 Resp: 3753
Ion Ratio Lower Upper
56 100
69 176.7 28.1 42.1#
84 143.4 66.8 100.2#




Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0






#33
1,2-Dichloroethane-d4
Concen: 52.148 ug/l
RT: 7.99 min Scan# 1933
Delta R.T. -0.00 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Tgt Ion: 65 Resp: 150363
Ion Ratio Lower Upper
65 100
67 53.0 0.0 101.4



Abundance Ion 64.90 (64.60 to 65.60): VN0
Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. -0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :

20200514-2801-B


Tgt Ion:114 Resp: 328978

Ion Ratio Lower Upper

114 100

63 22.9 0.0 47.6




88 16.5 0.0 34.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 49.238 ug/l

RT: 7.55 min Scan# 1796

Delta R.T. -0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59


Tgt Ion:113 Resp: 110804

Ion Ratio Lower Upper

113 100

111 102.7 81.8 122.6



192 19.3 15.0 22.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V









#36
 1,1-Dichloropropene
 Concen: 4.585 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN062957.D
 Acq: 15 Aug 2020 00:59




Instrument :
 MSVOA_N
 ClientSampleId :
 20200514-2801-B


Tgt Ion: 75 Resp: 14686
 Ion Ratio Lower Upper
 75 100
 110 5.5 15.6 46.7#
 77 0.0 24.1 36.1#



#37
 Ethyl Acetate
 Concen: 3.432 ug/l
 RT: 6.80 min Scan# 1563
 Delta R.T. -0.09 min
 Lab File: VN062957.D
 Acq: 15 Aug 2020 00:59

Tgt Ion: 43 Resp: 11485
 Ion Ratio Lower Upper
 43 100
 61 7.5 12.1 18.1#
 70 0.0 8.3 12.5#





#38

Carbon Tetrachloride

Concen: 4.707 ug/l

RT: 7.63 min Scan# 1821

Delta R.T. -0.11 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :

20200514-2801-B


Tgt Ion:117 Resp: 17833

Ion Ratio Lower Upper

117 100

119 6.8 79.4 119.2#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.240 ug/l

RT: 7.16 min Scan# 1676

Delta R.T. 0.03 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Tgt Ion: 41 Resp: 362


Ion Ratio Lower Upper

41 100

39 0.0 69.4 104.2#

67 0.0 0.0 0.0

52 0.0 0.0 0.0






Abundance Ion 41.00 (40.70 to 41.70): V

Ion 39.00 (38.70 to 39.70): V

Ion 67.00 (66.70 to 67.70): V

Ion 52.00 (51.70 to 52.70): V









#43
Isopropyl Acetate
Concen: 5.121 ug/l
RT: 8.13 min Scan# 1977
Delta R.T. 0.00 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59




Instrument :
MSVOA_N
ClientSampleId :
20200514-2801-B


Tgt Ion: 43 Resp: 28583
Ion Ratio Lower Upper
43 100
61 17.8 15.8 23.6
87 11.3 9.9 14.9



#48
Methyl methacrylate
Concen: 1.593 ug/l
RT: 9.15 min Scan# 2294
Delta R.T. -0.02 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Tgt Ion: 41 Resp: 4374
Ion Ratio Lower Upper
41 100
69 0.0 67.4 101.2#
39 21.0 47.6 71.4#





#50

Toluene-d8

Concen: 50.476 ug/l

RT: 10.06 min Scan# 2578

Delta R.T. -0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :


20200514-2801-B

Tgt Ion: 98 Resp: 443468




Ion Ratio Lower Upper

98 100

100 63.4 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VN062756.D (-2562) (-)



#51

4-Methyl-2-Pentanone

Concen: 2.419 ug/l

RT: 9.95 min Scan# 2544

Delta R.T. -0.00 min

Lab File: VN062957.D


Acq: 15 Aug 2020 00:59

Tgt Ion: 43 Resp: 8735



Ion Ratio Lower Upper


43 100

58 33.5 31.1 46.7



Abundance Ion 43.00 (42.70 to 43.70): VN062756.D (-2522) (-)









#54
 cis-1,3-Dichloropropene
 Concen: 4.260 ug/l
 RT: 9.88 min Scan# 2520
 Delta R.T. 0.07 min
 Lab File: VN062957.D
 Acq: 15 Aug 2020 00:59




Instrument :
 MSVOA_N
 ClientSampleId :
 20200514-2801-B


Tgt Ion: 75 Resp: 17911
 Ion Ratio Lower Upper
 75 100
 77 0.0 25.5 38.3#
 39 40.3 41.9 62.9#



#56
 Ethyl methacrylate
 Concen: 0.490 ug/l
 RT: 10.52 min Scan# 2719
 Delta R.T. 0.11 min
 Lab File: VN062957.D
 Acq: 15 Aug 2020 00:59

Tgt Ion: 69 Resp: 1860
 Ion Ratio Lower Upper
 69 100
 41 2561.2 52.2 78.2#
 39 1091.2 30.9 46.3#









#59
2-Hexanone
Concen: 4.323 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.01 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59




Instrument :
MSVOA_N
ClientSampleId :
20200514-2801-B


Tgt Ion: 43 Resp: 11644
Ion Ratio Lower Upper
43 100
58 22.6 26.7 80.1#



#62
4-Bromofluorobenzene
Concen: 49.727 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. -0.00 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Tgt Ion: 95 Resp: 165101
Ion Ratio Lower Upper
95 100
174 76.1 0.0 144.8
176 72.1 0.0 142.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. -0.00 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59




Instrument :
MSVOA_N
ClientSampled :
20200514-2801-B


Tgt Ion:117 Resp: 355466
Ion Ratio Lower Upper
117 100
82 59.3 49.0 73.4
119 31.7 26.2 39.2



#71
Bromoform
Concen: 0.523 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.27 min
Lab File: VN062957.D
Acq: 15 Aug 2020 00:59

Tgt Ion:173 Resp: 1276
Ion Ratio Lower Upper
173 100
175 638.6 24.0 72.0#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. -0.00 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :

20200514-2801-B


Tgt Ion:152 Resp: 125183

Ion Ratio Lower Upper

152 100

115 61.6 31.6 94.8




150 160.9 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.426 ug/l

RT: 12.00 min Scan# 3182

Delta R.T. -0.04 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Tgt Ion: 43 Resp: 1628


Ion Ratio Lower Upper

43 100

70 0.0 32.3 48.5#

55 12.5 20.2 30.2#

61 0.0 18.2 27.4#






Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 70.00 (69.70 to 70.70): VN0

Ion 55.00 (54.70 to 55.70): VN0

Ion 61.00 (60.70 to 61.70): VN0





#76

1,2,3-Trichloropropane

Concen: 24.743 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. -0.15 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :


20200514-2801-B

Tgt Ion: 75 Resp: 83534

Ion Ratio Lower Upper




75 100

77 1.1 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 77.00 (76.70 to 77.70): VN0



#81

trans-1,4-Dichloro-2-butene

Concen: 75.534 ug/l

RT: 12.38 min Scan# 3298

Delta R.T. 0.10 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59


Tgt Ion: 75 Resp: 83534

Ion Ratio Lower Upper

75 100

53 0.1 79.1 118.7#



89 0.0 36.2 54.2#




Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0





#92

1,2-Dibromo-3-Chloropropane

Concen: 0.400 ug/l

RT: 14.29 min Scan# 3893

Delta R.T. 0.04 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59

Instrument :

MSVOA_N

ClientSampleId :

20200514-2801-B


Tgt Ion: 75 Resp: 285

Ion Ratio Lower Upper

75 100

155 0.0 35.0 105.1#

157 0.0 43.2 129.6#




Abundance Ion 74.90 (74.60 to 75.60): VN062756.D (-3866) (-)

Ion 154.80 (154.50 to 155.50): VN062756.D (-3866) (-)

Ion 156.80 (156.50 to 157.50): VN062756.D (-3866) (-)

Time--> 14.26 14.28 14.30




Abundance Ion 74.90 (74.60 to 75.60): VN062957.D (-3786) (-)

Ion 154.80 (154.50 to 155.50): VN062957.D (-3786) (-)

Ion 156.80 (156.50 to 157.50): VN062957.D (-3786) (-)

Time--> 14.26 14.28 14.30



#95

Naphthalene

Concen: 3.713 ug/l

RT: 15.12 min Scan# 4151

Delta R.T. 0.02 min

Lab File: VN062957.D

Acq: 15 Aug 2020 00:59


Tgt Ion: 128 Resp: 29

Ion Ratio Lower Upper

128 100

127 0.0 10.8 16.2#

129 0.0 8.8 13.2#




Abundance Ion 128.00 (127.70 to 128.70): VN062756.D (-4132) (-)

Ion 127.00 (126.70 to 127.70): VN062756.D (-4132) (-)

Ion 129.00 (128.70 to 129.70): VN062756.D (-4132) (-)

Time--> 15.11 15.12 15.13



Abundance Ion 128.00 (127.70 to 128.70): VN062957.D (-4053) (-)

Ion 127.00 (126.70 to 127.70): VN062957.D (-4053) (-)

Ion 129.00 (128.70 to 129.70): VN062957.D (-4053) (-)

Time--> 15.11 15.12 15.13