

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062935.D
 Acq On : 14 Aug 2020 12:41
 Operator : JC/MD
 Sample : VN0814WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814WBL01

Quant Time: Aug 14 14:47:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 14 06:20:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	25221	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	122991	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	115986	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	70587	31.33	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	104.43%
60) 4-Bromofluorobenzene	12.38	95	45748	22.40	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	74.67%
63) Toluene-d8	10.06	98	154932	28.57	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	95.23%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062935.D
Acq On : 14 Aug 2020 12:41
Operator : JC/MD
Sample : VN0814WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814WBL01

Quant Time: Aug 14 14:47:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 14 06:20:27 2020
Response via : Initial Calibration

