

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062936.D
 Acq On : 14 Aug 2020 13:14
 Operator : JC/MD
 Sample : L3603-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD-(AUG)

Quant Time: Aug 14 14:48:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 14 06:20:27 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	29080	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	136579	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	128411	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	7.99	65	75518	29.07	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.90%
60) 4-Bromofluorobenzene	12.38	95	51824	22.92	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	76.40%
63) Toluene-d8	10.06	98	175061	29.15	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	97.17%

Target Compounds

					Ovalue	
15) Acetone	3.77	58	489023	2064.013	ug/l	95
30) 2-Butanone	6.79	43	16009	15.393	ug/l #	69
58) 4-Methyl-2-Pentanone	9.95	43	7349	3.289	ug/l #	89
62) Toluene	10.12	91	92120	13.606	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062936.D
Acq On : 14 Aug 2020 13:14
Operator : JC/MD
Sample : L3603-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD-(AUG)

Quant Time: Aug 14 14:48:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 14 06:20:27 2020
Response via : Initial Calibration

