

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062950.D
 Acq On : 14 Aug 2020 22:09
 Operator : JC/MD
 Sample : PB130954ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130954ZHE#09

Quant Time: Aug 17 10:04:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	198338	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	371358	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	397887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	137374	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	165789	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	7.55	113	126630	49.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.70%	
50) Toluene-d8	10.06	98	495680	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
62) 4-Bromofluorobenzene	12.38	95	176345	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
Target Compounds						
16) Acetone	3.78	43	40482	34.158	ug/l	97
18) Methyl Acetate	4.29	43	6605	1.255	ug/l #	72
20) Methylene Chloride	4.50	84	22990	9.992	ug/l	98
43) Isopropyl Acetate	8.13	43	75465	11.978	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062950.D
Acq On : 14 Aug 2020 22:09
Operator : JC/MD
Sample : PB130954ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130954ZHE#09

Quant Time: Aug 17 10:04:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

