

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062952.D
 Acq On : 14 Aug 2020 22:58
 Operator : JC/MD
 Sample : PB130954ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130954ZHE#11

Quant Time: Aug 17 10:09:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	189776	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	369056	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	378873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128915	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	157747	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
35) Dibromofluoromethane	7.55	113	121184	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
50) Toluene-d8	10.06	98	479333	48.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.26%	
62) 4-Bromofluorobenzene	12.38	95	164678	44.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.42%	
Target Compounds						
16) Acetone	3.78	43	38935	34.335	ug/l	99
18) Methyl Acetate	4.28	43	8401	2.049	ug/l #	83
20) Methylene Chloride	4.50	84	24068	11.005	ug/l	96
43) Isopropyl Acetate	8.13	43	71089	11.354	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062952.D
Acq On : 14 Aug 2020 22:58
Operator : JC/MD
Sample : PB130954ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130954ZHE#11

Quant Time: Aug 17 10:09:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

