

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062957.D
 Acq On : 15 Aug 2020 00:59
 Operator : JC/MD
 Sample : L3682-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200514-2801-B

Quant Time: Aug 17 10:21:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	174363	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	328978	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	355466	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125183	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	150363	52.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.30%	
35) Dibromofluoromethane	7.55	113	110804	49.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.48%	
50) Toluene-d8	10.06	98	443468	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	12.38	95	165101	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
Target Compounds						
16) Acetone	3.78	43	48699	46.741	ug/l	98
18) Methyl Acetate	4.29	43	7710	2.045	ug/l #	65
20) Methylene Chloride	4.50	84	15443	7.452	ug/l	96
25) 2-Butanone	6.80	43	11485	7.783	ug/l	96
43) Isopropyl Acetate	8.13	43	28583	5.121	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062957.D
Acq On : 15 Aug 2020 00:59
Operator : JC/MD
Sample : L3682-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200514-2801-B

Quant Time: Aug 17 10:21:47 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

