

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062958.D
 Acq On : 15 Aug 2020 1:23
 Operator : JC/MD
 Sample : L3659-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP1-1

Quant Time: Aug 17 10:23:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	347617	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	321782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134240	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	149517	54.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.80%	
35) Dibromofluoromethane	7.55	113	106564	44.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.62%	
50) Toluene-d8	10.06	98	469814	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	12.38	95	164781	46.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.94%	
Target Compounds						
16) Acetone	3.78	43	40063	40.713	ug/l	97
20) Methylene Chloride	4.51	84	7628	3.529	ug/l	91
43) Isopropyl Acetate	8.13	43	34288	5.814	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062958.D
Acq On : 15 Aug 2020 1:23
Operator : JC/MD
Sample : L3659-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP1-1

Quant Time: Aug 17 10:23:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

