

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062960.D
 Acq On : 15 Aug 2020 2:11
 Operator : JC/MD
 Sample : L3659-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP1-3

Quant Time: Aug 17 10:27:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	137744	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	249170	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	348873	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	123127	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	116143	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
35) Dibromofluoromethane	7.55	113	85345	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
50) Toluene-d8	10.06	98	359007	53.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.90%	
62) 4-Bromofluorobenzene	12.38	95	172683	68.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	137.34%	
Target Compounds						
16) Acetone	3.78	43	12459	15.137	ug/l	99
18) Methyl Acetate	4.29	43	4366	1.139	ug/l #	74
20) Methylene Chloride	4.50	84	8079	4.674	ug/l #	85
43) Isopropyl Acetate	8.13	43	45637	10.796	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062960.D
Acq On : 15 Aug 2020 2:11
Operator : JC/MD
Sample : L3659-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP1-3

Quant Time: Aug 17 10:27:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

