

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062961.D
 Acq On : 15 Aug 2020 2:36
 Operator : JC/MD
 Sample : L3508-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-081220

Quant Time: Aug 17 10:31:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	182682	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	347833	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	247522	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	89379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	155702	51.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.08%	
35) Dibromofluoromethane	7.55	113	121893	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
50) Toluene-d8	10.06	98	467222	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.38	95	113532	32.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.68%	
Target Compounds						
16) Acetone	3.78	43	22705	20.800	ug/l	94
18) Methyl Acetate	4.29	43	8405	2.175	ug/l	91
20) Methylene Chloride	4.50	84	12384	5.523	ug/l	98
43) Isopropyl Acetate	8.13	43	76085	12.893	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062961.D
Acq On : 15 Aug 2020 2:36
Operator : JC/MD
Sample : L3508-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-081220

Quant Time: Aug 17 10:31:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

