

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\  
 Data File : VN062962.D  
 Acq On : 15 Aug 2020 3:00  
 Operator : JC/MD  
 Sample : L3501-05  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 33 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 NB-08-081220

Quant Time: Aug 17 10:33:13 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N080520W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Aug 05 13:26:42 2020  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|--------|---------|----------|
| 1) Pentafluorobenzene       | 7.63  | 168  | 182268   | 50.00  | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.55  | 114  | 341455   | 50.00  | ug/l    | 0.00     |
| 63) Chlorobenzene-d5        | 11.38 | 117  | 248578   | 50.00  | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.32 | 152  | 86846    | 50.00  | ug/l    | 0.00     |
| System Monitoring Compounds |       |      |          |        |         |          |
| 33) 1,2-Dichloroethane-d4   | 7.99  | 65   | 153238   | 50.84  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 101.68% |          |
| 35) Dibromofluoromethane    | 7.55  | 113  | 116190   | 49.74  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 99.48%  |          |
| 50) Toluene-d8              | 10.06 | 98   | 308608   | 33.84  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 67.68%  |          |
| 62) 4-Bromofluorobenzene    | 12.38 | 95   | 110904   | 32.18  | ug/l    | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 64.36%  |          |
| Target Compounds            |       |      |          |        |         |          |
| 16) Acetone                 | 3.78  | 43   | 26103    | 23.967 | ug/l    | 93       |
| 18) Methyl Acetate          | 4.29  | 43   | 9439     | 2.593  | ug/l #  | 80       |
| 20) Methylene Chloride      | 4.51  | 84   | 11188    | 4.928  | ug/l #  | 95       |
| 43) Isopropyl Acetate       | 8.13  | 43   | 75987    | 13.117 | ug/l    | 99       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\  
Data File : VN062962.D  
Acq On : 15 Aug 2020 3:00  
Operator : JC/MD  
Sample : L3501-05  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 33 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
NB-08-081220

Quant Time: Aug 17 10:33:13 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_N\METHODS\82N080520W.M  
Quant Title : SW846 8260  
QLast Update : Wed Aug 05 13:26:42 2020  
Response via : Initial Calibration

