

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062964.D
 Acq On : 15 Aug 2020 3:48
 Operator : JC/MD
 Sample : L3682-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200516-2801-D

Quant Time: Aug 17 10:37:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	129871	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	236545	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	248049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	91544	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	109452	50.96	ug/l	0.00
Spiked Amount			Recovery	=	101.92%	
35) Dibromofluoromethane	7.55	113	82977	51.28	ug/l	0.00
Spiked Amount			Recovery	=	102.56%	
50) Toluene-d8	10.06	98	311992	49.39	ug/l	0.00
Spiked Amount			Recovery	=	98.78%	
62) 4-Bromofluorobenzene	12.38	95	110811	46.42	ug/l	0.00
Spiked Amount			Recovery	=	92.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	63341	81.622	ug/l	98
18) Methyl Acetate	4.28	43	8679	3.682	ug/l	91
20) Methylene Chloride	4.50	84	16366	10.930	ug/l	94
25) 2-Butanone	6.80	43	13272	12.075	ug/l	90
43) Isopropyl Acetate	8.13	43	40919	10.196	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062964.D
Acq On : 15 Aug 2020 3:48
Operator : JC/MD
Sample : L3682-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200516-2801-D

Quant Time: Aug 17 10:37:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

