

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN081420\
 Data File : VN062965.D
 Acq On : 15 Aug 2020 4:12
 Operator : JC/MD
 Sample : L3682-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 20200517-2801-E

Quant Time: Aug 17 10:39:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 05 13:26:42 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	131318	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	236680	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	250460	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	88152	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	110646	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
35) Dibromofluoromethane	7.55	113	81782	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
50) Toluene-d8	10.06	98	329364	52.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.22%	
62) 4-Bromofluorobenzene	12.38	95	112862	47.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.50%	
Target Compounds						
16) Acetone	3.78	43	39328	50.120	ug/l	99
20) Methylene Chloride	4.50	84	12338	7.953	ug/l	97
25) 2-Butanone	6.80	43	7570	6.812	ug/l	97
43) Isopropyl Acetate	8.13	43	33532	8.351	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN081420\
Data File : VN062965.D
Acq On : 15 Aug 2020 4:12
Operator : JC/MD
Sample : L3682-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
20200517-2801-E

Quant Time: Aug 17 10:39:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N080520W.M
Quant Title : SW846 8260
QLast Update : Wed Aug 05 13:26:42 2020
Response via : Initial Calibration

