

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081420\
 Data File : VN062934.D
 Acq On : 14 Aug 2020 11:35
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Aug 14 14:45:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 14 06:20:27 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Bromochloromethane	1.000	1.000	0.0	98	0.00
2 M	Dichlorodifluoromethane	2.572	2.687	-4.5	95	0.00
3 M	Chloromethane	2.228	1.990	10.7	94	0.00
4 M	Vinyl Chloride	2.354	2.037	13.5	96	0.00
5 M	Bromomethane	2.030	2.183	-7.5	105	0.00
6 M	Chloroethane	2.041	2.199	-7.7	103	0.00
7 M	Trichlorofluoromethane	4.589	4.647	-1.3	89	0.00
8 T	Diethyl Ether	1.004	0.940	6.4	95	0.00
9	1,1,2-Trichlorotrifluoroeth	1.510	1.540	-2.0	102	0.00
10 M	1,1-Dichloroethene	1.384	1.331	3.8	98	0.00
11	Methyl Iodide	1.922	1.776	7.6	97	0.00
12	Methyl Acetate	2.096	1.956	6.7	95	0.00
13 M	Acrolein	0.268	0.204	23.9	86	0.00
14 M	Acrylonitrile	0.915	0.834	8.9	95	0.00
15 M	Acetone	0.244	0.236	3.3	94	0.00
16 M	Carbon Disulfide	3.948	3.562	9.8	98	0.00
17	Allyl chloride	2.384	2.183	8.4	98	0.00
18 M	Methylene Chloride	1.791	1.724	3.7	99	0.00
19 M	trans-1,2-Dichloroethene	1.546	1.473	4.7	100	0.00
20 T	Diisopropyl ether	5.582	5.252	5.9	94	0.00
21 M	1,1-Dichloroethane	3.267	3.084	5.6	93	0.00
22 M	cis-1,2-Dichloroethene	1.848	1.843	0.3	101	0.00
23 M	tert-Butyl Alcohol	0.392	0.352	10.2	90	0.00
24 M	Methyl tert-Butyl Ether	6.015	5.583	7.2	94	0.00
25 M	Chloroform	3.545	3.578	-0.9	101	0.00
26	Cyclohexane	2.766	2.592	6.3	98	0.00
27 s	1,2-Dichloroethane-d4	2.680	2.697	-0.6	96	0.00
28 I	1,4-Difluorobenzene	1.000	1.000	0.0	96	0.00
29	1,1-Dichloropropene	0.443	0.449	-1.4	99	0.00
30 M	2-Butanone	0.228	0.219	3.9	95	0.00
31	2,2-Dichloropropane	0.441	0.551	-24.9	119	0.00
32 M	1,1,1-Trichloroethane	0.606	0.623	-2.8	100	0.00
33 M	Carbon Tetrachloride	0.538	0.564	-4.8	102	0.00
34 M	Benzene	1.330	1.326	0.3	98	0.00
35	Methacrylonitrile	0.203	0.198	2.5	109	0.00
36 M	1,2-Dichloroethane	0.578	0.582	-0.7	97	0.00
37 M	Trichloroethene	0.330	0.324	1.8	95	0.00
38	Methylcyclohexane	0.529	0.522	1.3	99	0.00
39 M	1,2-Dichloropropane	0.363	0.359	1.1	97	0.00
40	Dibromomethane	0.246	0.258	-4.9	101	0.00
41 M	Bromodichloromethane	0.552	0.566	-2.5	99	0.00
42 M	Vinyl Acetate	0.772	0.705	8.7	94	0.00
43	Ethyl Acetate	0.449	0.430	4.2	97	0.00
44	Isopropyl Acetate	0.733	0.687	6.3	97	0.00
45 T	1,4-Dioxane	0.007	0.007#	0.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081420\
 Data File : VN062934.D
 Acq On : 14 Aug 2020 11:35
 Operator : JC/MD
 Sample : VSTDCCC020
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC020

Quant Time: Aug 14 14:45:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Aug 14 06:20:27 2020
 Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46	Methyl methacrylate	0.366	0.334	8.7	94	0.00
47	n-amyl Acetate	0.654	0.571	12.7	91	0.00
48 M	t-1,3-Dichloropropene	0.577	0.563	2.4	97	0.00
49 T	cis-1,3-Dichloropropene	0.580	0.589	-1.6	102	0.00
50 M	1,1,2-Trichloroethane	0.354	0.365	-3.1	100	0.00
51	Ethyl methacrylate	0.538	0.502	6.7	97	0.00
52	1,3-Dichloropropane	0.611	0.605	1.0	96	0.00
53 M	Dibromochloromethane	0.423	0.442	-4.5	102	0.00
54 M	1,2-Dibromoethane	0.362	0.370	-2.2	102	0.00
55 M	2-Chloroethyl vinyl ether	0.130	0.110	15.4	98	0.00
56 M	Bromoform	0.300	0.300	0.0	99	0.00
57 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
58 M	4-Methyl-2-Pentanone	0.522	0.479	8.2	94	0.00
59 M	2-Hexanone	0.381	0.344	9.7	93	0.00
60 S	4-Bromofluorobenzene	0.528	0.511	3.2	101	0.00
61 M	Tetrachloroethene	0.310	0.318	-2.6	106	0.00
62 M	Toluene	1.582	1.525	3.6	98	0.00
63 S	Toluene-d8	1.403	1.405	-0.1	99	0.00
64 M	Chlorobenzene	0.992	0.959	3.3	100	0.00
65	1,1,1,2-Tetrachloroethane	0.410	0.396	3.4	97	0.00
66 M	Ethyl Benzene	1.761	1.682	4.5	101	0.00
67 M	m/p-Xylenes	0.652	0.654	-0.3	102	0.00
68 M	o-Xylene	0.626	0.596	4.8	100	0.00
69 M	Styrene	1.111	1.066	4.1	99	0.00
70	Isopropylbenzene	1.781	1.725	3.1	101	0.00
71 M	1,1,2,2-Tetrachloroethane	0.586	0.570	2.7	100	0.00
72	1,2,3-Trichloropropane	0.580	0.555	4.3	93	0.00
73	Bromobenzene	0.437	0.438	-0.2	105	0.00
74	n-propylbenzene	2.055	1.980	3.6	103	0.00
75	2-Chlorotoluene	1.290	1.247	3.3	100	0.00
76	1,3,5-Trimethylbenzene	1.583	1.540	2.7	103	0.00
77	t-1,4-Dichloro-2-butene	0.187	0.169	9.6	100	0.00
78	4-Chlorotoluene	1.312	1.262	3.8	102	0.00
79	tert-butylbenzene	1.308	1.242	5.0	101	0.00
80	1,2,4-Trimethylbenzene	1.547	1.461	5.6	101	0.00
81	sec-Butylbenzene	1.727	1.633	5.4	102	0.00
82	p-Isopropyltoluene	1.544	1.477	4.3	104	0.00
83 M	1,3-Dichlorobenzene	0.765	0.746	2.5	104	0.00
84 M	1,4-Dichlorobenzene	0.751	0.732	2.5	107	0.00
85	n-Butylbenzene	1.316	1.092	17.0	103	0.00
86 T	Hexachloroethane	0.334	0.310	7.2	104	0.00
87 M	1,2-Dichlorobenzene	0.790	0.718	9.1	102	0.00
88	1,2-Dibromo-3-Chloropropane	0.138	0.124	10.1	113	0.00
89	1,2,4-Trichlorobenzene	0.364	0.313	14.0	110	0.00
90	Hexachlorobutadiene	0.243	0.269	-10.7	124	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN081420\
Data File : VN062934.D
Acq On : 14 Aug 2020 11:35
Operator : JC/MD
Sample : VSTDCCC020
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC020

Quant Time: Aug 14 14:45:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N081420W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Aug 14 06:20:27 2020
Response via : Initial Calibration

Min. RRF : 0.050 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
91 M	Naphthalene	0.985	0.723	26.6	102	0.00
92	1,2,3-Trichlorobenzene	0.364	0.317	12.9	111	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0