

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083296.D
 Acq On : 14 Aug 2024 13:16
 Operator : JC\MD
 Sample : VN0814MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0814MBL01

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 08/15/2024
 Supervised By :Semsettin Yesilyurt 08/15/2024

Quant Time: Aug 14 23:21:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.871	168	132429m	50.000	ug/l	-0.35
34) 1,4-Difluorobenzene	8.859	114	262359	50.000	ug/l	-0.24
63) Chlorobenzene-d5	11.812	117	265516m	50.000	ug/l	-0.05
72) 1,4-Dichlorobenzene-d4	13.782	152	118344	50.000	ug/l	-0.01

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.259	65	107726	57.150	ug/l	-0.32
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.300%
35) Dibromofluoromethane	7.789	113	85354	52.122	ug/l	-0.38
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.240%
50) Toluene-d8	10.447	98	324818	53.174	ug/l	-0.12
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.340%
62) 4-Bromofluorobenzene	12.824	95	133119m	55.898	ug/l	-0.02
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083296.D
Acq On : 14 Aug 2024 13:16
Operator : JC\MD
Sample : VN0814MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0814MBL01

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 08/15/2024
Supervised By :Semsettin Yesilyurt 08/15/2024

Quant Time: Aug 14 23:21:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

