

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083309.D
 Acq On : 14 Aug 2024 18:29
 Operator : JC\MD
 Sample : P3596-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 918-J-WS-081324

Quant Time: Aug 15 00:37:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	253276	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	263485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	116254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	108126	59.126	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.260%	
35) Dibromofluoromethane	8.171	113	86138	54.487	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	108.980%	
50) Toluene-d8	10.571	98	317935	53.914	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	107.820%	
62) 4-Bromofluorobenzene	12.847	95	133097	57.893	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	115.780%	
Target Compounds						
16) Acetone	4.442	43	3479	4.862	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083309.D
Acq On : 14 Aug 2024 18:29
Operator : JC\MD
Sample : P3596-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
918-J-WS-081324

Quant Time: Aug 15 00:37:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

