

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083311.D
 Acq On : 14 Aug 2024 19:18
 Operator : JC\MD
 Sample : P3601-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 918-J-WPO-0.25-081324

Quant Time: Aug 14 23:25:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	132738	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	257654	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	268755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	120062	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	108236	57.287	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.580%			
35) Dibromofluoromethane	8.165	113	86279	53.649	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.300%			
50) Toluene-d8	10.565	98	330454	55.084	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.160%			
62) 4-Bromofluorobenzene	12.853	95	136247	58.256	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 116.520%			
Target Compounds						
16) Acetone	4.442	43	3433	4.644	ug/l	96
29) Tetrahydrofuran	7.853	42	7284	9.926	ug/l	91
52) Toluene	10.629	92	15081	3.293	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083311.D
Acq On : 14 Aug 2024 19:18
Operator : JC\MD
Sample : P3601-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
918-J-WPO-0.25-081324

Quant Time: Aug 14 23:25:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

