

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
 Data File : VN083312.D
 Acq On : 14 Aug 2024 19:42
 Operator : JC\MD
 Sample : P3601-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 918-J-WPO-0.25-081324-FD

Quant Time: Aug 14 23:26:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	125717	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	247320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	259031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117034	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	104805	58.569	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 117.140%	
35) Dibromofluoromethane	8.171	113	83752	54.253	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.500%	
50) Toluene-d8	10.571	98	309735	53.788	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.580%	
62) 4-Bromofluorobenzene	12.847	95	127869	56.958	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 113.920%	
Target Compounds						Qvalue
16) Acetone	4.436	43	2896	4.136	ug/l	# 79
29) Tetrahydrofuran	7.841	42	5451	7.843	ug/l	# 85
52) Toluene	10.630	92	14433	3.284	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN081424\
Data File : VN083312.D
Acq On : 14 Aug 2024 19:42
Operator : JC\MD
Sample : P3601-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
918-J-WPO-0.25-081324-FD

Quant Time: Aug 14 23:26:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

